

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085786.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0219WBSD01	Chloromethane	20	16.5	ug/L	83	4		50	139	20
	Vinyl chloride	20	16.8	ug/L	84	2		58	137	20
	Bromomethane	20	17.1	ug/L	86	0		53	141	20
	Chloroethane	20	16.2	ug/L	81	5		60	138	20
	Trichlorofluoromethane	20	16.7	ug/L	84	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.1	ug/L	86	3		70	136	20
	1,1-Dichloroethene	20	17.2	ug/L	86	1		71	131	20
	Acetone	100	90.3	ug/L	90	4		39	160	20
	Carbon disulfide	20	15.9	ug/L	79	6		64	133	20
	Methyl tert-butyl Ether	20	19.2	ug/L	96	6		71	124	20
	Methylene Chloride	20	17.2	ug/L	86	2		74	124	20
	trans-1,2-Dichloroethene	20	16.9	ug/L	85	0		75	124	20
	1,1-Dichloroethane	20	17.3	ug/L	86	3		77	125	20
	2-Butanone	100	91.4	ug/L	91	1		56	143	20
	Carbon Tetrachloride	20	17.0	ug/L	85	3		72	136	20
	cis-1,2-Dichloroethene	20	17.6	ug/L	88	0		78	123	20
	Chloroform	20	17.5	ug/L	88	0		79	124	20
	1,1,1-Trichloroethane	20	17.2	ug/L	86	2		74	131	20
	Methylcyclohexane	20	17.3	ug/L	86	9		72	132	20
	Benzene	20	17.8	ug/L	89	2		79	120	20
	1,2-Dichloroethane	20	18.2	ug/L	91	2		73	128	20
	Trichloroethene	20	16.7	ug/L	84	2		79	123	20
	1,2-Dichloropropane	20	18.1	ug/L	91	2		78	122	20
	Bromodichloromethane	20	18.0	ug/L	90	2		79	125	20
	4-Methyl-2-Pentanone	100	96.3	ug/L	96	8		67	130	20
	Toluene	20	18.8	ug/L	94	1		80	121	20
	t-1,3-Dichloropropene	20	18.7	ug/L	94	1		73	127	20
	cis-1,3-Dichloropropene	20	18.8	ug/L	94	1		75	124	20
	1,1,2-Trichloroethane	20	17.9	ug/L	90	1		80	119	20
	2-Hexanone	100	97.6	ug/L	98	4		57	139	20
	Dibromochloromethane	20	18.9	ug/L	95	4		74	126	20
	Tetrachloroethene	20	17.5	ug/L	88	2		74	129	20
	Chlorobenzene	20	17.5	ug/L	88	2		82	118	20
	Ethyl Benzene	20	17.9	ug/L	90	4		79	121	20
	m/p-Xylenes	40	37.6	ug/L	94	3		80	121	20
	o-Xylene	20	18.5	ug/L	93	3		78	122	20
	Styrene	20	18.1	ug/L	91	3		78	123	20
	Bromoform	20	18.5	ug/L	93	1		66	130	20
	Isopropylbenzene	20	18.1	ug/L	91	3		72	131	20
	1,1,2,2-Tetrachloroethane	20	17.6	ug/L	88	6		71	121	20
	1,3-Dichlorobenzene	20	17.4	ug/L	87	2		80	119	20
	1,4-Dichlorobenzene	20	17.1	ug/L	86	2		79	118	20
	1,2-Dichlorobenzene	20	17.1	ug/L	86	2		80	119	20