

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085782.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_N Time Analyzed: 10:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	338951	8.22	542050	9.09	475442	11.86
UPPER LIMIT	677902	8.718	1084100	9.594	950884	12.359
LOWER LIMIT	169476	7.718	271025	8.594	237721	11.359
EPA SAMPLE NO.						
BP-VPB-192-EB-20250212	292071	8.22	537683	9.10	478583	11.86
BP-VPB-192-TB-20250210	264567	8.22	487878	9.10	434147	11.86
BP-VPB-192-GW-780-782	259651	8.22	470979	9.10	403744	11.87
VPB192-HYD-20250214	297765	8.22	540496	9.09	473230	11.86
VN0219WBL01	286037	8.22	514740	9.09	457806	11.86
VN0219WBS01	317605	8.22	521252	9.09	448364	11.86
VN0219WBSD01	291949	8.22	482343	9.09	419097	11.86

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.