

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021236.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_Y Time Analyzed: 14:55
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	200462	7.72	311138	8.63	266562	11.43
UPPER LIMIT	400924	8.219	622276	9.128	533124	11.926
LOWER LIMIT	100231	7.219	155569	8.128	133281	10.926
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727	223938	7.71	424683	8.62	405408	11.42
VY0219SBL01	252953	7.72	478949	8.62	429832	11.43
VY0219SBS01	187422	7.71	302089	8.62	255752	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.