

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085782.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_N Time Analyzed: 10:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	234508	13.788				
UPPER LIMIT	469016	14.288				
LOWER LIMIT	117254	13.288				
EPA SAMPLE NO.						
BP-VPB-192-EB-20250212	200119	13.79				
BP-VPB-192-TB-20250210	174960	13.79				
BP-VPB-192-GW-780-782	157449	13.79				
VPB192-HYD-20250214	194311	13.79				
VN0219WBL01	185142	13.79				
VN0219WBS01	214627	13.79				
VN0219WBSD01	199880	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.