

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
 Lab File ID: VY021320.D Date Analyzed: 02/26/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:11
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	131935	13.347				
UPPER LIMIT	263870	13.847				
LOWER LIMIT	65967.5	12.847				
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727MS	76053	13.35				
BP-VPB-192-GW-725-727MSD	63404 *	13.35				
VY0226SBL01	192201	13.35				
VY0226SBS01	128725	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.