

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134760

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP84585,MP84586,MP84587,MP84588,MP84589,MP84590
ICV Standard	MP84591
CCV Standard	MP84593
ICSA Standard	
CRI Standard	MP84595
LCS Standard	
Chk Standard	MP84592,MP84594,MP84596,MP84598

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/20/25 12:05		Mohan	OK
2	S0.2	S0.2	CAL2	02/20/25 12:07		Mohan	OK
3	S2.5	S2.5	CAL3	02/20/25 12:12		Mohan	OK
4	S5	S5	CAL4	02/20/25 12:14		Mohan	OK
5	S7.5	S7.5	CAL5	02/20/25 12:17		Mohan	OK
6	S10	S10	CAL6	02/20/25 12:19		Mohan	OK
7	ICV11	ICV11	ICV	02/20/25 12:23		Mohan	OK
8	ICB11	ICB11	ICB	02/20/25 12:28		Mohan	OK
9	CCV91	CCV91	CCV	02/20/25 12:30		Mohan	OK
10	CCB91	CCB91	CCB	02/20/25 12:35		Mohan	OK
11	CRA	CRA	CRDL	02/20/25 12:46		Mohan	OK
12	HighStd	HighStd	HIGH STD	02/20/25 12:49		Mohan	OK
13	ChkStd	ChkStd	SAM	02/20/25 12:51		Mohan	OK
14	PB166807BL	PB166807BL	MB	02/20/25 12:56		Mohan	OK
15	PB166807BS	PB166807BS	LCS	02/20/25 12:58		Mohan	OK
16	Q1390-02	3772	SAM	02/20/25 13:00		Mohan	OK
17	Q1393-02	SOIL-COMP	SAM	02/20/25 13:03		Mohan	OK
18	Q1393-02DUP	SOIL-COMPDUP	DUP	02/20/25 13:05		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134760

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP84585,MP84586,MP84587,MP84588,MP84589,MP84590
ICV Standard	MP84591
CCV Standard	MP84593
ICSA Standard	
CRI Standard	MP84595
LCS Standard	
Chk Standard	MP84592,MP84594,MP84596,MP84598

19	Q1393-02MS	SOIL-COMPMS	MS	02/20/25 13:07		Mohan	OK
20	Q1393-02MSD	SOIL-COMPMSD	MSD	02/20/25 13:10		Mohan	OK
21	CCV92	CCV92	CCV	02/20/25 13:12		Mohan	OK
22	CCB92	CCB92	CCB	02/20/25 13:17		Mohan	OK
23	Q1393-04	OD-281	SAM	02/20/25 13:19		Mohan	OK
24	PB166808BL	PB166808BL	MB	02/20/25 13:22		Mohan	OK
25	PB166808BS	PB166808BS	LCS	02/20/25 13:24		Mohan	OK
26	Q1381-01	OUTFALL-001	SAM	02/20/25 13:26		Mohan	OK
27	Q1381-02	OUTFALL-002	SAM	02/20/25 13:28		Mohan	OK
28	Q1390-01	3771	SAM	02/20/25 13:31		Mohan	OK
29	Q1392-01	435	SAM	02/20/25 13:33		Mohan	OK
30	Q1393-05	PURGE-WATER-COM	SAM	02/20/25 13:35		Mohan	OK
31	Q1393-05DUP	PURGE-WATER-COM	DUP	02/20/25 13:37		Mohan	OK
32	Q1393-05MS	PURGE-WATER-COM	MS	02/20/25 13:40		Mohan	OK
33	CCV93	CCV93	CCV	02/20/25 13:42		Mohan	OK
34	CCB93	CCB93	CCB	02/20/25 13:47		Mohan	OK
35	Q1393-05MSD	PURGE-WATER-COM	MSD	02/20/25 13:49		Mohan	OK
36	PB166784TB	PB166784TB	MB	02/20/25 13:54		Mohan	OK
37	Q1393-02L	SOIL-COMPL	SD	02/20/25 13:57		Mohan	OK
38	Q1393-02A	SOIL-COMPA	PS	02/20/25 13:59		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134760

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP84585,MP84586,MP84587,MP84588,MP84589,MP84590
ICV Standard	MP84591
CCV Standard	MP84593
ICSA Standard	
CRI Standard	MP84595
LCS Standard	
Chk Standard	MP84592,MP84594,MP84596,MP84598

39	Q1393-05L	PURGE-WATER-COM	SD	02/20/25 14:01		Mohan	OK
40	Q1393-05A	PURGE-WATER-COM	PS	02/20/25 14:04		Mohan	OK
41	CCV94	CCV94	CCV	02/20/25 14:06		Mohan	OK
42	CCB94	CCB94	CCB	02/20/25 14:08		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134892

Review By	jaswal	Review On	3/5/2025 8:49:15 AM
Supervise By	Mohan	Supervise On	3/5/2025 4:19:12 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84642		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84642		
LCS Standard			
Chk Standard	MP84649,MP84650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/04/25 09:42		Kareem	OK
2	S1	S1	CAL2	03/04/25 09:46		Kareem	OK
3	S2	S2	CAL3	03/04/25 09:51		Kareem	OK
4	S3	S3	CAL4	03/04/25 09:55		Kareem	OK
5	S4	S4	CAL5	03/04/25 09:59		Kareem	OK
6	S5	S5	CAL6	03/04/25 10:03		Kareem	OK
7	ICV01	ICV01	ICV	03/04/25 10:07	ICV Fail for Ba, K, and Na, (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	03/04/25 10:13		Kareem	OK
9	ICB01	ICB01	ICB	03/04/25 10:17		Kareem	OK
10	CRI01	CRI01	CRDL	03/04/25 10:22		Kareem	OK
11	ICSA01	ICSA01	ICSA	03/04/25 10:26		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	03/04/25 10:30		Kareem	OK
13	ICSADL	ICSADL	ICSA	03/04/25 10:34		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	03/04/25 10:39		Kareem	OK
15	CCV01	CCV01	CCV	03/04/25 10:43		Kareem	OK
16	CCB01	CCB01	CCB	03/04/25 10:47		Kareem	OK
17	Q1398-06	BUS-DISCONNECT	SAM	03/04/25 10:51		Kareem	OK
18	Q1398-07	HANDHOLD	SAM	03/04/25 10:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134892

Review By	jaswal	Review On	3/5/2025 8:49:15 AM
Supervise By	Mohan	Supervise On	3/5/2025 4:19:12 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84642		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84642		
LCS Standard			
Chk Standard	MP84649,MP84650		

19	Q1453-01	MOO-25-0061	SAM	03/04/25 11:00		Kareem	OK
20	Q1454-03DL	LAW-25-0043DL	SAM	03/04/25 11:05	Not required	Kareem	Not Ok
21	Q1458-01	SP-SOIL	SAM	03/04/25 11:09		Kareem	OK
22	Q1458-03	SP-SOIL	SAM	03/04/25 11:13		Kareem	OK
23	Q1456-02	EFF-WASTE WATER	SAM	03/04/25 11:17		Kareem	OK
24	Q1383-01	OU4-PCS-TC-11-021	SAM	03/04/25 11:22		Kareem	OK
25	Q1383-03	OU4-PCS-TC-12-021	SAM	03/04/25 11:26		Kareem	OK
26	Q1383-05	OU4-PCS-TC-13-021	SAM	03/04/25 11:30		Kareem	OK
27	CCV02	CCV02	CCV	03/04/25 11:34		Kareem	OK
28	CCB02	CCB02	CCB	03/04/25 11:39		Kareem	OK
29	Q1383-07	OU4-PCS-TC-14-021	SAM	03/04/25 11:43		Kareem	OK
30	Q1383-09	OU4-PCS-TC-15-021	SAM	03/04/25 11:47		Kareem	OK
31	Q1383-11	OU4-PCS-TC-16-021	SAM	03/04/25 11:52		Kareem	OK
32	Q1383-13	OU4-PCS-TC-17-021	SAM	03/04/25 11:56		Kareem	OK
33	Q1383-13DUP	OU4-PCS-TC-17-021	DUP	03/04/25 12:00		Kareem	OK
34	Q1383-13L	OU4-PCS-TC-17-021	SD	03/04/25 12:04		Kareem	OK
35	Q1383-13MS	OU4-PCS-TC-17-021	MS	03/04/25 12:08	0.1ML EACH M6003, M6011	Kareem	OK
36	Q1383-13MSD	OU4-PCS-TC-17-021	MSD	03/04/25 12:13	0.1ML EACH M6003, M6011	Kareem	OK
37	Q1383-13A	OU4-PCS-TC-17-021	PS	03/04/25 12:17	0.1ML EACH M6003, M6011	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134892

Review By	jaswal	Review On	3/5/2025 8:49:15 AM
Supervise By	Mohan	Supervise On	3/5/2025 4:19:12 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84642
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84642
LCS Standard	
Chk Standard	MP84649,MP84650

38	Q1383-15	OU4-PCS-TC-18-021	SAM	03/04/25 12:21		Kareem	OK
39	CCV03	CCV03	CCV	03/04/25 12:26		Kareem	OK
40	CCB03	CCB03	CCB	03/04/25 12:30		Kareem	OK
41	Q1383-17	OU4-PCS-TC-19-021	SAM	03/04/25 12:34		Kareem	OK
42	Q1383-19	OU4-PCS-TC-20-021	SAM	03/04/25 12:38		Kareem	OK
43	Q1383-21	OU4-CF-15-021725	SAM	03/04/25 12:43		Kareem	OK
44	PB166768BL	PB166768BL	MB	03/04/25 12:52		Kareem	OK
45	PB166768BS	PB166768BS	LCS	03/04/25 12:56		Kareem	OK
46	Q1381-01	OUTFALL-001	SAM	03/04/25 13:00		Kareem	OK
47	Q1381-02	OUTFALL-002	SAM	03/04/25 13:04		Kareem	OK
48	Q1381-02DUP	OUTFALL-002DUP	DUP	03/04/25 13:08		Kareem	OK
49	Q1381-02L	OUTFALL-002L	SD	03/04/25 13:13		Kareem	OK
50	Q1381-02MS	OUTFALL-002MS	MS	03/04/25 13:17	0.1ML EACH M6003, M6011	Kareem	OK
51	CCV04	CCV04	CCV	03/04/25 13:27		Kareem	OK
52	CCB04	CCB04	CCB	03/04/25 13:31		Kareem	OK
53	PB166777BL	PB166777BL	MB	03/04/25 13:44		Kareem	OK
54	LR1	LR1	HIGH STD	03/04/25 13:49		Kareem	OK
55	LR2	LR2	HIGH STD	03/04/25 13:53		Kareem	OK
56	Q1381-02MSD	OUTFALL-002MSD	MSD	03/04/25 13:58	0.1ML EACH M6003, M6011	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134892

Review By	jaswal	Review On	3/5/2025 8:49:15 AM
Supervise By	Mohan	Supervise On	3/5/2025 4:19:12 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84642
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84642
LCS Standard	
Chk Standard	MP84649,MP84650

57	Q1381-02A	OUTFALL-002A	PS	03/04/25 14:02	0.1ML EACH M6003, M6011	Kareem	OK
58	CCV05	CCV05	CCV	03/04/25 14:08		Kareem	OK
59	CCB05	CCB05	CCB	03/04/25 14:12		Kareem	OK
60	PB166777BS	PB166777BS	LCS	03/04/25 14:16		Kareem	OK
61	Q1454-03	LAW-25-0043	SAM	03/04/25 14:20		Kareem	OK
62	CCV06	CCV06	CCV	03/04/25 14:30		Kareem	OK
63	CCB06	CCB06	CCB	03/04/25 14:34		Kareem	OK