



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	RSB Environmental				SDG No.:	Q1381				
Contract:	RSBE01	Lab Code:	CHEM		Case No.:	Q1381	SAS No.:	Q1381		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB11	Mercury	0.20	+/-0.20	U	0.20	CV	02/20/2025	12:28	LB134760	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RSB Environmental

SDG No.: Q1381

Contract: RSBE01

Lab Code: CHEM

Case No.: Q1381

SAS No.: Q1381

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB91	Mercury	0.20	+/-0.20	U	0.20	CV	02/20/2025	12:35	LB134760
CCB92	Mercury	0.20	+/-0.20	U	0.20	CV	02/20/2025	13:17	LB134760
CCB93	Mercury	0.20	+/-0.20	U	0.20	CV	02/20/2025	13:47	LB134760
CCB94	Mercury	0.20	+/-0.20	U	0.20	CV	02/20/2025	14:08	LB134760

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>Q1381</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1381</u>	SAS No.: <u>Q1381</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/04/2025	10:17	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	10:17	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	10:17	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	10:17	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	10:17	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	10:17	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	10:17	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	10:17	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	10:17	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	10:17	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	10:17	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	10:17	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	10:17	LB134892

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>Q1381</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1381</u>	SAS No.: <u>Q1381</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/04/2025	10:47	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	10:47	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	10:47	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	10:47	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	10:47	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	10:47	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	10:47	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	10:47	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	10:47	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	10:47	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	10:47	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	10:47	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	10:47	LB134892
CCB02	Aluminum	100	+/-100	U	100	P	03/04/2025	11:39	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	11:39	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	11:39	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	11:39	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	11:39	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	11:39	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	11:39	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	11:39	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	11:39	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	11:39	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	11:39	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	11:39	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	11:39	LB134892
CCB03	Aluminum	100	+/-100	U	100	P	03/04/2025	12:30	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	12:30	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	12:30	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	12:30	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	12:30	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	12:30	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	12:30	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	12:30	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	12:30	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	12:30	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	12:30	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	12:30	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	12:30	LB134892

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>Q1381</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1381</u>	SAS No.: <u>Q1381</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Aluminum	100	+/-100	U	100	P	03/04/2025	13:31	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	13:31	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	13:31	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	13:31	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	13:31	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	13:31	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	13:31	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	13:31	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	13:31	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	13:31	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	13:31	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	13:31	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	13:31	LB134892
CCB05	Aluminum	100	+/-100	U	100	P	03/04/2025	14:12	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	14:12	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	14:12	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	14:12	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	14:12	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	14:12	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	14:12	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	14:12	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	14:12	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	14:12	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	14:12	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	14:12	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	14:12	LB134892
CCB06	Aluminum	100	+/-100	U	100	P	03/04/2025	14:34	LB134892
	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	14:34	LB134892
	Barium	100	+/-100	U	100	P	03/04/2025	14:34	LB134892
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	14:34	LB134892
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	14:34	LB134892
	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	14:34	LB134892
	Iron	100	+/-100	U	100	P	03/04/2025	14:34	LB134892
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	14:34	LB134892
	Manganese	20.0	+/-20.0	U	20.0	P	03/04/2025	14:34	LB134892
	Nickel	40.0	+/-40.0	U	40.0	P	03/04/2025	14:34	LB134892
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	14:34	LB134892
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	14:34	LB134892
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	14:34	LB134892