

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RSB Environmental

SDG No.: Q1381

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166808BL		WATER		Batch Number:	PB166808		Prep Date:	02/20/2025	
	Mercury	0.20	<0.20	U	0.20	CV	02/20/2025	13:22	LB134760

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RSB Environmental

SDG No.: Q1381

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166777BL	WATER			Batch Number:	PB166777		Prep Date:	02/19/2025	
	Aluminum	50.0	<50.0	U	50.0	P	03/04/2025	13:44	LB134892
	Arsenic	10.0	<10.0	U	10.0	P	03/04/2025	13:44	LB134892
	Barium	50.0	<50.0	U	50.0	P	03/04/2025	13:44	LB134892
	Cadmium	3.00	<3.00	U	3.00	P	03/04/2025	13:44	LB134892
	Chromium	5.00	<5.00	U	5.00	P	03/04/2025	13:44	LB134892
	Copper	10.0	<10.0	U	10.0	P	03/04/2025	13:44	LB134892
	Iron	50.0	<50.0	U	50.0	P	03/04/2025	13:44	LB134892
	Lead	6.00	<6.00	U	6.00	P	03/04/2025	13:44	LB134892
	Manganese	10.0	<10.0	U	10.0	P	03/04/2025	13:44	LB134892
	Nickel	20.0	<20.0	U	20.0	P	03/04/2025	13:44	LB134892
	Selenium	10.0	<10.0	U	10.0	P	03/04/2025	13:44	LB134892
	Silver	5.00	<5.00	U	5.00	P	03/04/2025	13:44	LB134892
	Zinc	20.0	<20.0	U	20.0	P	03/04/2025	13:44	LB134892