

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RSB Environmental **SDG No.:** Q1381
Contract: RSBE01 **Lab Code:** CHEM **Case No.:** Q1381 **SAS No.:** Q1381

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166777BS							
Aluminum	ug/L	1000	927		93	80 - 120	P
Arsenic	ug/L	400	374		94	80 - 120	P
Barium	ug/L	100	84.4		84	80 - 120	P
Cadmium	ug/L	100	86.7		87	80 - 120	P
Chromium	ug/L	200	207		104	80 - 120	P
Copper	ug/L	150	147		98	80 - 120	P
Iron	ug/L	1500	1430		95	80 - 120	P
Lead	ug/L	500	439		88	80 - 120	P
Manganese	ug/L	100	91.5		92	80 - 120	P
Nickel	ug/L	250	230		92	80 - 120	P
Selenium	ug/L	1000	881		88	80 - 120	P
Silver	ug/L	37.5	34.2		91	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RSB Environmental

SDG No.: Q1381

Contract: RSBE01

Lab Code: CHEM

Case No.: Q1381

SAS No.: Q1381

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166808BS Mercury	ug/L	4.0	3.55		89	80 - 120	CV