

Initial and Continuing Calibration Blank Summary

Client: Nobis Group

SDG No.: Q1382

Project: Raymark Superfund Site

RunNo.: LB134755

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025
Sample ID: CCB4 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025
Sample ID: CCB5 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/19/2025