



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	88.6	100	89	40 - 160	P	02/19/2025	17:20	LB134756
	Antimony	52.1	50.0	104	40 - 160	P	02/19/2025	17:20	LB134756
	Arsenic	19.1	20.0	96	40 - 160	P	02/19/2025	17:20	LB134756
	Barium	102	100	102	40 - 160	P	02/19/2025	17:20	LB134756
	Beryllium	5.48	6.0	91	40 - 160	P	02/19/2025	17:20	LB134756
	Cadmium	5.92	6.0	99	40 - 160	P	02/19/2025	17:20	LB134756
	Calcium	1930	2000	96	40 - 160	P	02/19/2025	17:20	LB134756
	Chromium	9.54	10.0	95	40 - 160	P	02/19/2025	17:20	LB134756
	Cobalt	29.4	30.0	98	40 - 160	P	02/19/2025	17:20	LB134756
	Copper	22.1	20.0	111	40 - 160	P	02/19/2025	17:20	LB134756
	Iron	102	100	102	40 - 160	P	02/19/2025	17:20	LB134756
	Lead	11.7	12.0	97	40 - 160	P	02/19/2025	17:20	LB134756
	Magnesium	2000	2000	100	40 - 160	P	02/19/2025	17:20	LB134756
	Manganese	20.4	20.0	102	40 - 160	P	02/19/2025	17:20	LB134756
	Nickel	39.6	40.0	99	40 - 160	P	02/19/2025	17:20	LB134756
	Potassium	2110	2000	106	40 - 160	P	02/19/2025	17:20	LB134756
	Selenium	19.6	20.0	98	40 - 160	P	02/19/2025	17:20	LB134756
	Silver	10.1	10.0	101	40 - 160	P	02/19/2025	17:20	LB134756
	Sodium	2140	2000	107	40 - 160	P	02/19/2025	17:20	LB134756
	Thallium	39.0	40.0	98	40 - 160	P	02/19/2025	17:20	LB134756
	Vanadium	39.3	40.0	98	40 - 160	P	02/19/2025	17:20	LB134756
	Zinc	40.7	40.0	102	40 - 160	P	02/19/2025	17:20	LB134756
CRA	Mercury	0.21	0.2	105	40 - 160	CV	02/21/2025	13:04	LB134771