



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Nobis Group				SDG No.:	Q1382				
Contract:	NOBI03	Lab Code:	CHEM		Case No.:	Q1382	SAS No.:	Q1382		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB13	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	12:54	LB134771	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>							
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	13:01	LB134771
CCB02	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	13:32	LB134771
CCB03	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	14:02	LB134771
CCB04	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	14:33	LB134771
CCB05	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	15:06	LB134771
CCB06	Mercury	0.20	+/-0.20	U	0.20	CV	02/21/2025	15:24	LB134771

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>							
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	02/19/2025	17:15	LB134756
	Antimony	50.0	+/-50.0	U	50.0	P	02/19/2025	17:15	LB134756
	Arsenic	20.0	+/-20.0	U	20.0	P	02/19/2025	17:15	LB134756
	Barium	100	+/-100	U	100	P	02/19/2025	17:15	LB134756
	Beryllium	6.00	+/-6.00	U	6.00	P	02/19/2025	17:15	LB134756
	Cadmium	6.00	+/-6.00	U	6.00	P	02/19/2025	17:15	LB134756
	Calcium	2000	+/-2000	U	2000	P	02/19/2025	17:15	LB134756
	Chromium	10.0	+/-10.0	U	10.0	P	02/19/2025	17:15	LB134756
	Cobalt	30.0	+/-30.0	U	30.0	P	02/19/2025	17:15	LB134756
	Copper	20.0	+/-20.0	U	20.0	P	02/19/2025	17:15	LB134756
	Iron	100	+/-100	U	100	P	02/19/2025	17:15	LB134756
	Lead	12.0	+/-12.0	U	12.0	P	02/19/2025	17:15	LB134756
	Magnesium	2000	+/-2000	U	2000	P	02/19/2025	17:15	LB134756
	Manganese	20.0	+/-20.0	U	20.0	P	02/19/2025	17:15	LB134756
	Nickel	40.0	+/-40.0	U	40.0	P	02/19/2025	17:15	LB134756
	Potassium	2000	+/-2000	U	2000	P	02/19/2025	17:15	LB134756
	Selenium	20.0	+/-20.0	U	20.0	P	02/19/2025	17:15	LB134756
	Silver	10.0	+/-10.0	U	10.0	P	02/19/2025	17:15	LB134756
	Sodium	2000	+/-2000	U	2000	P	02/19/2025	17:15	LB134756
	Thallium	40.0	+/-40.0	U	40.0	P	02/19/2025	17:15	LB134756
	Vanadium	40.0	+/-40.0	U	40.0	P	02/19/2025	17:15	LB134756
	Zinc	40.0	+/-40.0	U	40.0	P	02/19/2025	17:15	LB134756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>							
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	02/19/2025	17:36	LB134756
	Antimony	50.0	+/-50.0	U	50.0	P	02/19/2025	17:36	LB134756
	Arsenic	20.0	+/-20.0	U	20.0	P	02/19/2025	17:36	LB134756
	Barium	100	+/-100	U	100	P	02/19/2025	17:36	LB134756
	Beryllium	6.00	+/-6.00	U	6.00	P	02/19/2025	17:36	LB134756
	Cadmium	6.00	+/-6.00	U	6.00	P	02/19/2025	17:36	LB134756
	Calcium	2000	+/-2000	U	2000	P	02/19/2025	17:36	LB134756
	Chromium	10.0	+/-10.0	U	10.0	P	02/19/2025	17:36	LB134756
	Cobalt	30.0	+/-30.0	U	30.0	P	02/19/2025	17:36	LB134756
	Copper	20.0	+/-20.0	U	20.0	P	02/19/2025	17:36	LB134756
	Iron	100	+/-100	U	100	P	02/19/2025	17:36	LB134756
	Lead	12.0	+/-12.0	U	12.0	P	02/19/2025	17:36	LB134756
	Magnesium	2000	+/-2000	U	2000	P	02/19/2025	17:36	LB134756
	Manganese	20.0	+/-20.0	U	20.0	P	02/19/2025	17:36	LB134756
	Nickel	40.0	+/-40.0	U	40.0	P	02/19/2025	17:36	LB134756
	Potassium	2000	+/-2000	U	2000	P	02/19/2025	17:36	LB134756
	Selenium	20.0	+/-20.0	U	20.0	P	02/19/2025	17:36	LB134756
	Silver	10.0	+/-10.0	U	10.0	P	02/19/2025	17:36	LB134756
	Sodium	2000	+/-2000	U	2000	P	02/19/2025	17:36	LB134756
	Thallium	40.0	+/-40.0	U	40.0	P	02/19/2025	17:36	LB134756
	Vanadium	40.0	+/-40.0	U	40.0	P	02/19/2025	17:36	LB134756
	Zinc	40.0	+/-40.0	U	40.0	P	02/19/2025	17:36	LB134756
CCB02	Aluminum	100	+/-100	U	100	P	02/19/2025	18:27	LB134756
	Antimony	50.0	+/-50.0	U	50.0	P	02/19/2025	18:27	LB134756
	Arsenic	20.0	+/-20.0	U	20.0	P	02/19/2025	18:27	LB134756
	Barium	100	+/-100	U	100	P	02/19/2025	18:27	LB134756
	Beryllium	6.00	+/-6.00	U	6.00	P	02/19/2025	18:27	LB134756
	Cadmium	6.00	+/-6.00	U	6.00	P	02/19/2025	18:27	LB134756
	Calcium	2000	+/-2000	U	2000	P	02/19/2025	18:27	LB134756
	Chromium	10.0	+/-10.0	U	10.0	P	02/19/2025	18:27	LB134756
	Cobalt	30.0	+/-30.0	U	30.0	P	02/19/2025	18:27	LB134756
	Copper	20.0	+/-20.0	U	20.0	P	02/19/2025	18:27	LB134756
	Iron	100	+/-100	U	100	P	02/19/2025	18:27	LB134756
	Lead	12.0	+/-12.0	U	12.0	P	02/19/2025	18:27	LB134756
	Magnesium	2000	+/-2000	U	2000	P	02/19/2025	18:27	LB134756
	Manganese	20.0	+/-20.0	U	20.0	P	02/19/2025	18:27	LB134756
	Nickel	40.0	+/-40.0	U	40.0	P	02/19/2025	18:27	LB134756
	Potassium	2000	+/-2000	U	2000	P	02/19/2025	18:27	LB134756
	Selenium	20.0	+/-20.0	U	20.0	P	02/19/2025	18:27	LB134756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>							
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	02/19/2025	18:27	LB134756
	Sodium	2000	+/-2000	U	2000	P	02/19/2025	18:27	LB134756
	Thallium	40.0	+/-40.0	U	40.0	P	02/19/2025	18:27	LB134756
	Vanadium	40.0	+/-40.0	U	40.0	P	02/19/2025	18:27	LB134756
	Zinc	40.0	+/-40.0	U	40.0	P	02/19/2025	18:27	LB134756
CCB03	Aluminum	100	+/-100	U	100	P	02/19/2025	19:13	LB134756
	Antimony	50.0	+/-50.0	U	50.0	P	02/19/2025	19:13	LB134756
	Arsenic	20.0	+/-20.0	U	20.0	P	02/19/2025	19:13	LB134756
	Barium	100	+/-100	U	100	P	02/19/2025	19:13	LB134756
	Beryllium	6.00	+/-6.00	U	6.00	P	02/19/2025	19:13	LB134756
	Cadmium	6.00	+/-6.00	U	6.00	P	02/19/2025	19:13	LB134756
	Calcium	2000	+/-2000	U	2000	P	02/19/2025	19:13	LB134756
	Chromium	10.0	+/-10.0	U	10.0	P	02/19/2025	19:13	LB134756
	Cobalt	30.0	+/-30.0	U	30.0	P	02/19/2025	19:13	LB134756
	Copper	20.0	+/-20.0	U	20.0	P	02/19/2025	19:13	LB134756
	Iron	100	+/-100	U	100	P	02/19/2025	19:13	LB134756
	Lead	12.0	+/-12.0	U	12.0	P	02/19/2025	19:13	LB134756
	Magnesium	2000	+/-2000	U	2000	P	02/19/2025	19:13	LB134756
	Manganese	20.0	+/-20.0	U	20.0	P	02/19/2025	19:13	LB134756
	Nickel	40.0	+/-40.0	U	40.0	P	02/19/2025	19:13	LB134756
	Potassium	2000	+/-2000	U	2000	P	02/19/2025	19:13	LB134756
	Selenium	20.0	+/-20.0	U	20.0	P	02/19/2025	19:13	LB134756
	Silver	10.0	+/-10.0	U	10.0	P	02/19/2025	19:13	LB134756
	Sodium	2000	+/-2000	U	2000	P	02/19/2025	19:13	LB134756
	Thallium	40.0	+/-40.0	U	40.0	P	02/19/2025	19:13	LB134756
	Vanadium	40.0	+/-40.0	U	40.0	P	02/19/2025	19:13	LB134756
	Zinc	40.0	+/-40.0	U	40.0	P	02/19/2025	19:13	LB134756
CCB04	Aluminum	100	+/-100	U	100	P	02/19/2025	19:34	LB134756
	Antimony	50.0	+/-50.0	U	50.0	P	02/19/2025	19:34	LB134756
	Arsenic	20.0	+/-20.0	U	20.0	P	02/19/2025	19:34	LB134756
	Barium	100	+/-100	U	100	P	02/19/2025	19:34	LB134756
	Beryllium	6.00	+/-6.00	U	6.00	P	02/19/2025	19:34	LB134756
	Cadmium	6.00	+/-6.00	U	6.00	P	02/19/2025	19:34	LB134756
	Calcium	2000	+/-2000	U	2000	P	02/19/2025	19:34	LB134756
	Chromium	10.0	+/-10.0	U	10.0	P	02/19/2025	19:34	LB134756
	Cobalt	30.0	+/-30.0	U	30.0	P	02/19/2025	19:34	LB134756
	Copper	20.0	+/-20.0	U	20.0	P	02/19/2025	19:34	LB134756
	Iron	100	+/-100	U	100	P	02/19/2025	19:34	LB134756
	Lead	12.0	+/-12.0	U	12.0	P	02/19/2025	19:34	LB134756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>							
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	02/19/2025	19:34	LB134756
	Manganese	20.0	+/-20.0	U	20.0	P	02/19/2025	19:34	LB134756
	Nickel	40.0	+/-40.0	U	40.0	P	02/19/2025	19:34	LB134756
	Potassium	2000	+/-2000	U	2000	P	02/19/2025	19:34	LB134756
	Selenium	20.0	+/-20.0	U	20.0	P	02/19/2025	19:34	LB134756
	Silver	10.0	+/-10.0	U	10.0	P	02/19/2025	19:34	LB134756
	Sodium	2000	+/-2000	U	2000	P	02/19/2025	19:34	LB134756
	Thallium	40.0	+/-40.0	U	40.0	P	02/19/2025	19:34	LB134756
	Vanadium	40.0	+/-40.0	U	40.0	P	02/19/2025	19:34	LB134756
	Zinc	40.0	+/-40.0	U	40.0	P	02/19/2025	19:34	LB134756