

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166760BS							
Aluminum	mg/Kg	88.9	83.4		94	74 - 119	P
Antimony	mg/Kg	35.6	34.9		98	79 - 114	P
Arsenic	mg/Kg	35.6	35.8		101	82 - 111	P
Barium	mg/Kg	8.9	8.65		97	83 - 113	P
Beryllium	mg/Kg	8.9	8.31		93	83 - 113	P
Cadmium	mg/Kg	8.9	8.71		98	82 - 113	P
Calcium	mg/Kg	44.4	42.9	J	97	81 - 116	P
Chromium	mg/Kg	17.8	18.0		101	85 - 113	P
Cobalt	mg/Kg	8.9	8.72		98	85 - 112	P
Copper	mg/Kg	13.3	13.7		103	81 - 117	P
Iron	mg/Kg	130	136		105	81 - 118	P
Lead	mg/Kg	44.4	43.8		99	81 - 112	P
Magnesium	mg/Kg	88.9	82.5	J	93	78 - 115	P
Manganese	mg/Kg	8.9	8.65		97	84 - 114	P
Nickel	mg/Kg	22.2	22.0		99	83 - 113	P
Potassium	mg/Kg	440	444		101	81 - 116	P
Selenium	mg/Kg	88.9	87.4		98	78 - 111	P
Silver	mg/Kg	3.3	3.37		102	82 - 112	P
Sodium	mg/Kg	130	137		105	83 - 118	P
Thallium	mg/Kg	88.9	89.5		101	83 - 111	P
Vanadium	mg/Kg	13.3	12.7		96	82 - 114	P
Zinc	mg/Kg	8.9	8.91		100	82 - 113	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166830BS Mercury	mg/Kg	0.25	0.24		95	80 - 124	CV