



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1382</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1382</u>	SAS No.:	<u>Q1382</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB14	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	10:36	LB134788

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>								
Contract: <u>NOBI03</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1382</u>			SAS No.: <u>Q1382</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	10:40	LB134788
CCB08	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	11:11	LB134788
CCB09	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	11:38	LB134788
CCB10	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	12:05	LB134788
CCB11	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	12:40	LB134788

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/25/2025	11:41	LB134821
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:41	LB134821
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/25/2025	11:41	LB134821
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:41	LB134821
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:41	LB134821
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/25/2025	11:41	LB134821
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/25/2025	11:41	LB134821
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/25/2025	11:41	LB134821
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:41	LB134821
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/25/2025	11:41	LB134821
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:41	LB134821
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:41	LB134821
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/25/2025	11:41	LB134821
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/25/2025	11:41	LB134821

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1382</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1382</u>	SAS No.: <u>Q1382</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/25/2025	11:55	LB134821
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:55	LB134821
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/25/2025	11:55	LB134821
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:55	LB134821
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:55	LB134821
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/25/2025	11:55	LB134821
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/25/2025	11:55	LB134821
	Lead	0.18	+/-1.00	J	0.75	1.00	P	02/25/2025	11:55	LB134821
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	11:55	LB134821
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/25/2025	11:55	LB134821
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:55	LB134821
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	11:55	LB134821
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/25/2025	11:55	LB134821
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/25/2025	11:55	LB134821
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/25/2025	12:40	LB134821
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	12:40	LB134821
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/25/2025	12:40	LB134821
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	12:40	LB134821
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	12:40	LB134821
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/25/2025	12:40	LB134821
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/25/2025	12:40	LB134821
	Lead	0.14	+/-1.00	J	0.75	1.00	P	02/25/2025	12:40	LB134821
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	12:40	LB134821
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/25/2025	12:40	LB134821
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	12:40	LB134821
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	12:40	LB134821
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/25/2025	12:40	LB134821
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/25/2025	12:40	LB134821
CCB03	Antimony	0.11	+/-2.00	J	0.25	2.00	P	02/25/2025	13:23	LB134821
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	13:23	LB134821
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/25/2025	13:23	LB134821
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	13:23	LB134821
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	13:23	LB134821
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/25/2025	13:23	LB134821
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/25/2025	13:23	LB134821
	Lead	0.16	+/-1.00	J	0.75	1.00	P	02/25/2025	13:23	LB134821
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/25/2025	13:23	LB134821
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/25/2025	13:23	LB134821
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	13:23	LB134821

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1382</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1382</u>	SAS No.:	<u>Q1382</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/25/2025	13:23	LB134821
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/25/2025	13:23	LB134821
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/25/2025	13:23	LB134821