

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1382

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166841BL		WATER		Batch Number:		PB166841		Prep Date:	02/24/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	02/24/2025	10:49	LB134788
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166841TB		WATER		Batch Number:		PB166841		Prep Date:	02/24/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	02/24/2025	12:17	LB134788

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1382

Instrument: P8

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166778BL	WATER			Batch Number:		PB166778		Prep Date:	02/19/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	02/25/2025	12:02	LB134821
	Arsenic	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:02	LB134821
	Barium	10.0	<10.0	U	1.25	10.0	P	02/25/2025	12:02	LB134821
	Beryllium	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:02	LB134821
	Cadmium	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:02	LB134821
	Chromium	2.00	<2.00	U	0.75	2.00	P	02/25/2025	12:02	LB134821
	Copper	2.00	<2.00	U	1.50	2.00	P	02/25/2025	12:02	LB134821
	Lead	1.00	<1.00	U	0.75	1.00	P	02/25/2025	12:02	LB134821
	Nickel	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:02	LB134821
	Selenium	5.00	<5.00	U	4.50	5.00	P	02/25/2025	12:02	LB134821
	Silver	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:02	LB134821
	Thallium	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:02	LB134821
	Vanadium	5.00	<5.00	U	0.25	5.00	P	02/25/2025	12:02	LB134821
	Zinc	5.00	<5.00	U	1.50	5.00	P	02/25/2025	12:02	LB134821
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166778TB	WATER			Batch Number:		PB166778		Prep Date:	02/19/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	02/25/2025	12:12	LB134821
	Arsenic	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:12	LB134821
	Barium	0.31	<10.0	J	1.25	10.0	P	02/25/2025	12:12	LB134821
	Beryllium	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:12	LB134821
	Cadmium	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:12	LB134821
	Chromium	2.00	<2.00	U	0.75	2.00	P	02/25/2025	12:12	LB134821
	Copper	2.00	<2.00	U	1.50	2.00	P	02/25/2025	12:12	LB134821
	Lead	0.12	<1.00	J	0.75	1.00	P	02/25/2025	12:12	LB134821
	Nickel	1.00	<1.00	U	0.25	1.00	P	02/25/2025	12:12	LB134821
	Selenium	5.00	<5.00	U	4.50	5.00	P	02/25/2025	12:12	LB134821
	Silver	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:12	LB134821
	Thallium	1.00	<1.00	U	0.50	1.00	P	02/25/2025	12:12	LB134821
	Vanadium	5.00	<5.00	U	0.25	5.00	P	02/25/2025	12:12	LB134821
	Zinc	5.00	<5.00	U	1.50	5.00	P	02/25/2025	12:12	LB134821