

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>Nobis Group</u>	SDG No.:	<u>Q1382</u>
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q1382</u>
		SAS No.:	<u>Q1382</u>
		Instrument ID:	<u>P8</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Antimony	1.08	1.5	72	-2.5	5.5	02/25/2025	11:45	LB134821
	Arsenic	0.31	0.1	310	-1.9	2.1	02/25/2025	11:45	LB134821
	Barium	1.52	1.2	127	-18.8	21.2	02/25/2025	11:45	LB134821
	Beryllium	0.31			-2	2	02/25/2025	11:45	LB134821
	Cadmium	0.40	0.7	57	-1.3	2.7	02/25/2025	11:45	LB134821
	Chromium	19.0	21.0	90	17	25	02/25/2025	11:45	LB134821
	Copper	8.03	8.0	100	4	12	02/25/2025	11:45	LB134821
	Lead	4.40	4.0	110	2	6	02/25/2025	11:45	LB134821
	Nickel	5.08	6.0	85	4	8	02/25/2025	11:45	LB134821
	Selenium	0.11	0.3	37	-9.7	10	02/25/2025	11:45	LB134821
	Silver	0.060			-2	2	02/25/2025	11:45	LB134821
	Thallium	0.13			-2	2	02/25/2025	11:45	LB134821
ICSAB01	Vanadium	0.15	0.5	30	-9.5	10.5	02/25/2025	11:45	LB134821
	Zinc	11.2	11.0	102	7	15	02/25/2025	11:45	LB134821
	Antimony	20.5	22.0	93	18	26	02/25/2025	11:48	LB134821
	Arsenic	20.8	19.0	110	16.2	21.9	02/25/2025	11:48	LB134821
	Barium	21.5	22.0	98	2	42	02/25/2025	11:48	LB134821
	Beryllium	20.4	19.0	107	16.2	21.9	02/25/2025	11:48	LB134821
	Cadmium	19.0	20.0	95	17	23	02/25/2025	11:48	LB134821
	Chromium	37.6	40.0	94	34	46	02/25/2025	11:48	LB134821
	Copper	27.6	25.0	110	21	29	02/25/2025	11:48	LB134821
	Lead	23.7	25.0	95	21.3	28.8	02/25/2025	11:48	LB134821
	Nickel	24.0	24.0	100	20.4	27.6	02/25/2025	11:48	LB134821
	Selenium	19.9	19.0	105	9	29	02/25/2025	11:48	LB134821
	Silver	18.1	18.0	101	15.3	20.7	02/25/2025	11:48	LB134821
	Thallium	19.9	21.0	95	17.9	24.2	02/25/2025	11:48	LB134821
	Vanadium	19.3	19.0	102	9	29	02/25/2025	11:48	LB134821
	Zinc	30.6	29.0	106	25	33	02/25/2025	11:48	LB134821