

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166778BS							
Antimony	ug/L	500	499		100	85 - 117	P
Arsenic	ug/L	500	505		101	84 - 116	P
Barium	ug/L	2500	2480		99	86 - 114	P
Beryllium	ug/L	500	501		100	83 - 121	P
Cadmium	ug/L	500	506		101	87 - 115	P
Chromium	ug/L	500	512		102	85 - 116	P
Copper	ug/L	5000	5130		103	85 - 118	P
Lead	ug/L	2500	2510		100	88 - 115	P
Nickel	ug/L	500	485		97	85 - 117	P
Selenium	ug/L	500	506		101	80 - 120	P
Silver	ug/L	500	505		101	85 - 116	P
Thallium	ug/L	500	501		100	82 - 116	P
Vanadium	ug/L	500	487		97	86 - 115	P
Zinc	ug/L	5000	5200		104	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1382
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1382 **SAS No.:** Q1382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166841BS Mercury	ug/L	4.0	4.37		109	82 - 119	CV