

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1382

Sas No.: Q1382 SDG No.: Q1382

Instrument ID: P8

Start Date : 02/25/2025

Run Number: LB134821

End Date : 02/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1100	100		100		100		100		100	
S2	S2	1103	101		100		101		99		100	
S3	S3	1107	98		96		98		97		99	
S4	S4	1110	97		97		98		94		98	
S5	S5	1113	96		96		97		93		98	
S6	S6	1116	96		97		97		94		99	
S7	S7	1118	97		99		99		93		100	
S8	S8	1121	88		92		92		81		92	
ICV01	ICV01	1133	99		99		99		98		98	
LLICV	LLICV	1136	99		99		99		98		98	
ICB01	ICB01	1141	99		99		98		98		99	
ICSA01	ICSA01	1145	94		95		96		88		97	
ICSAB01	ICSAB01	1148	93		94		96		88		95	
CCV01	CCV01	1151	89		93		92		83		92	
CCB01	CCB01	1155	98		97		98		97		97	
CRI	CRI	1158	97		98		97		98		98	
PB166778BL	PB166778BL	1202	96		96		96		97		96	
PB166778BS	PB166778BS	1207	94		94		94		90		96	
PB166778TB	PB166778TB	1212	96		95		97		96		95	
Q1382-02	OU4-PCS-TC-	1215	97		97		100		94		98	
Q1382-04	OU4-PCS-TC-	1218	98		99		104		96		98	
Q1382-06	OU4-PCS-TC-	1222	99		100		104		96		98	
Q1382-08	OU4-PCS-TC-	1225	99		99		102		95		98	
Q1382-10	OU4-PCS-TC-	1228	97		98		103		95		98	
Q1382-12	OU4-PCS-TC-	1231	98		100		105		96		98	
CCV02	CCV02	1236	90		94		93		86		95	
CCB02	CCB02	1240	97		98		97		96		96	
Q1382-14	OU4-PCS-TC-	1244	98		98		104		96		98	
Q1382-16	OU4-PCS-TC-	1248	98		101		105		97		99	
Q1382-18	OU4-PCS-TC-	1251	100		101		106		97		100	
Q1382-20	OU4-PCS-TC-	1254	99		101		106		98		99	
Q1382-20DUP	OU4-PCS-TC-	1257	98		101		106		97		100	
Q1382-20L	OU4-PCS-TC-	1301	98		100		101		98		98	
Q1382-20MS	OU4-PCS-TC-	1304	96		98		104		95		98	
Q1382-20MSD	OU4-PCS-TC-	1307	97		100		104		96		99	
Q1382-20A	OU4-PCS-TC-	1310	98		99		105		96		100	

Internal Standard %RI Limit: 30 - 120

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1382

Sas No.: Q1382 SDG No.: Q1382

Instrument ID: P8

Start Date : 02/25/2025

Run Number: LB134821

End Date : 02/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	ZZZZZZ	1316										
CCV03	CCV03	1320	89		92		92		83		93	
CCB03	CCB03	1323	96		97		97		96		96	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1382

Sas No.: Q1382 SDG No.: Q1382

Instrument ID: P8

Start Date : 02/25/2025

Run Number: LB134821

End Date : 02/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1100	100		100		100		100		100	
S2	S2	1103	100		98		99		98		98	
S3	S3	1107	97		96		97		97		98	
S4	S4	1110	96		95		95		98		98	
S5	S5	1113	97		97		94		97		98	
S6	S6	1116	98		97		95		98		98	
S7	S7	1118	102		100		94		99		101	
S8	S8	1121	96		96		83		93		94	
ICV01	ICV01	1133	99		99		98		98		99	
LLICV	LLICV	1136	99		99		99		99		100	
ICB01	ICB01	1141	98		97		99		98		98	
ICSA01	ICSA01	1145	96		95		89		96		96	
ICSAB01	ICSAB01	1148	97		96		90		96		96	
CCV01	CCV01	1151	96		94		86		94		94	
CCB01	CCB01	1155	98		100		99		99		99	
CRI	CRI	1158	97		98		98		98		99	
PB166778BL	PB166778BL	1202	95		95		96		95		97	
PB166778BS	PB166778BS	1207	97		96		92		95		97	
PB166778TB	PB166778TB	1212	95		96		97		97		98	
Q1382-02	OU4-PCS-TC-	1215	98		102		96		98		99	
Q1382-04	OU4-PCS-TC-	1218	99		104		97		98		98	
Q1382-06	OU4-PCS-TC-	1222	98		103		96		98		98	
Q1382-08	OU4-PCS-TC-	1225	98		103		96		98		99	
Q1382-10	OU4-PCS-TC-	1228	100		104		96		99		98	
Q1382-12	OU4-PCS-TC-	1231	100		105		98		99		99	
CCV02	CCV02	1236	98		96		87		94		95	
CCB02	CCB02	1240	97		97		99		97		97	
Q1382-14	OU4-PCS-TC-	1244	99		104		97		99		99	
Q1382-16	OU4-PCS-TC-	1248	100		105		98		100		100	
Q1382-18	OU4-PCS-TC-	1251	101		107		99		100		100	
Q1382-20	OU4-PCS-TC-	1254	102		108		99		100		100	
Q1382-20DUP	OU4-PCS-TC-	1257	101		105		98		99		99	
Q1382-20L	OU4-PCS-TC-	1301	100		100		99		99		99	
Q1382-20MS	OU4-PCS-TC-	1304	98		103		96		98		99	
Q1382-20MSD	OU4-PCS-TC-	1307	101		106		97		100		100	
Q1382-20A	OU4-PCS-TC-	1310	100		106		97		99		100	

Internal Standard %RI Limit: 30 - 120

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1382

Sas No.: Q1382 SDG No.: Q1382

Instrument ID: P8

Start Date : 02/25/2025

Run Number: LB134821

End Date : 02/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	ZZZZZZ	1316										
CCV03	CCV03	1320	94		94		85		93		95	
CCB03	CCB03	1323	97		98		97		97		98	