

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 02/19/2025 02/19/2025

Continuing Calib Time: 02:44 Initial Calibration Time(s): 10:21 11:57

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.34	7.34	7.24	7.44	0.00
2,4-DCAA	7.16	7.16	7.06	7.26	0.00
Dalapon	2.60	2.60	2.50	2.70	0.00
DICHLORPROP	8.04	8.04	7.94	8.14	0.00
2,4-D	8.27	8.27	8.17	8.37	0.00
2,4,5-TP(Silvex)	9.14	9.14	9.04	9.24	0.00
2,4,5-T	9.43	9.43	9.33	9.53	0.00
2,4-DB	10.00	10.00	9.90	10.10	0.00
Dinoseb	11.20	11.20	11.10	11.30	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 02/19/2025 02/19/2025

Continuing Calib Time: 02:44 Initial Calibration Time(s): 10:21 11:57

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.83	7.83	7.73	7.93	0.00
2,4-DCAA	7.64	7.64	7.54	7.74	0.00
Dalapon	2.65	2.65	2.55	2.75	0.00
DICHLORPROP	8.54	8.54	8.44	8.64	0.00
2,4-D	8.87	8.87	8.77	8.97	0.00
2,4,5-TP(Silvex)	9.76	9.77	9.67	9.87	0.01
2,4,5-T	10.18	10.18	10.08	10.28	0.00
2,4-DB	10.74	10.75	10.65	10.85	0.01
Dinoseb	11.12	11.12	11.02	11.22	0.00