

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/26/2025 Initial Calibration Date(s): 02/21/2025 02/21/2025

Continuing Calib Time: 19:23 Initial Calibration Time(s): 19:56 21:32

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.34	7.34	7.24	7.44	0.00
2,4-DCAA	7.16	7.16	7.06	7.26	0.00
Dalapon	2.60	2.60	2.50	2.70	0.00
DICHLORPROP	8.04	8.04	7.94	8.14	0.00
2,4-D	8.27	8.27	8.17	8.37	0.00
2,4,5-TP(Silvex)	9.14	9.14	9.04	9.24	0.00
2,4,5-T	9.43	9.43	9.33	9.53	0.00
2,4-DB	10.00	10.00	9.90	10.10	0.00
Dinoseb	11.19	11.19	11.09	11.29	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/26/2025 Initial Calibration Date(s): 02/21/2025 02/21/2025

Continuing Calib Time: 19:23 Initial Calibration Time(s): 19:56 21:32

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.82	7.83	7.73	7.93	0.01
2,4-DCAA	7.63	7.63	7.53	7.73	0.00
Dalapon	2.65	2.65	2.55	2.75	0.00
DICHLORPROP	8.53	8.54	8.44	8.64	0.01
2,4-D	8.86	8.86	8.76	8.96	0.00
2,4,5-TP(Silvex)	9.75	9.76	9.66	9.86	0.01
2,4,5-T	10.17	10.17	10.07	10.27	0.01
2,4-DB	10.73	10.73	10.63	10.83	0.00
Dinoseb	11.10	11.11	11.01	11.21	0.01