

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/19/2025 02/19/2025

Client Sample No.: CCAL01 Date Analyzed: 02/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029207.D Time Analyzed: 16:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.429	9.329	9.529	688.590	712.500	-3.4
2,4,5-TP(Silvex)	9.138	9.039	9.239	690.580	712.500	-3.1
2,4-D	8.270	8.170	8.370	664.580	705.000	-5.7
2,4-DB	9.997	9.898	10.098	669.260	712.500	-6.1
2,4-DCAA	7.158	7.059	7.259	708.710	750.000	-5.5
Dalapon	2.603	2.504	2.704	642.170	682.500	-5.9
DICAMBA	7.342	7.242	7.442	681.670	705.000	-3.3
DICHLORPROP	8.042	7.942	8.142	658.440	705.000	-6.6
Dinoseb	11.196	11.097	11.297	671.890	705.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/19/2025 02/19/2025

Client Sample No.: CCAL01 Date Analyzed: 02/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029207.D Time Analyzed: 16:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.178	10.081	10.281	700.300	712.500	-1.7
2,4,5-TP(Silvex)	9.763	9.666	9.866	703.290	712.500	-1.3
2,4-D	8.867	8.769	8.969	687.530	705.000	-2.5
2,4-DB	10.742	10.645	10.845	695.460	712.500	-2.4
2,4-DCAA	7.637	7.539	7.739	729.430	750.000	-2.7
Dalapon	2.650	2.553	2.753	656.470	682.500	-3.8
DICAMBA	7.832	7.734	7.934	698.370	705.000	-0.9
DICHLORPROP	8.542	8.444	8.644	681.970	705.000	-3.3
Dinoseb	11.119	11.021	11.221	681.670	705.000	-3.3