

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/19/2025 02/19/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029219.D Time Analyzed: 21:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.428	9.329	9.529	653.740	712.500	-8.2
2,4,5-TP(Silvex)	9.138	9.039	9.239	658.510	712.500	-7.6
2,4-D	8.270	8.170	8.370	613.270	705.000	-13.0
2,4-DB	9.997	9.898	10.098	610.610	712.500	-14.3
2,4-DCAA	7.158	7.059	7.259	638.970	750.000	-14.8
Dalapon	2.605	2.504	2.704	611.430	682.500	-10.4
DICAMBA	7.341	7.242	7.442	637.490	705.000	-9.6
DICHLORPROP	8.041	7.942	8.142	604.590	705.000	-14.2
Dinoseb	11.195	11.097	11.297	631.960	705.000	-10.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/19/2025 02/19/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029219.D Time Analyzed: 21:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.177	10.081	10.281	691.270	712.500	-3.0
2,4,5-TP(Silvex)	9.762	9.666	9.866	697.420	712.500	-2.1
2,4-D	8.866	8.769	8.969	678.860	705.000	-3.7
2,4-DB	10.740	10.645	10.845	678.560	712.500	-4.8
2,4-DCAA	7.636	7.539	7.739	722.690	750.000	-3.6
Dalapon	2.650	2.553	2.753	650.540	682.500	-4.7
DICAMBA	7.831	7.734	7.934	692.790	705.000	-1.7
DICHLORPROP	8.541	8.444	8.644	679.640	705.000	-3.6
Dinoseb	11.117	11.021	11.221	669.930	705.000	-5.0