

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/26/2025

Lab Sample No.: HSTDCCC750 Data File : PS029291.D Time Analyzed: 12:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.432	9.329	9.529	761.190	712.500	6.8
2,4,5-TP(Silvex)	9.141	9.038	9.238	770.220	712.500	8.1
2,4-D	8.274	8.171	8.371	753.460	705.000	6.9
2,4-DB	10.001	9.898	10.098	781.280	712.500	9.7
2,4-DCAA	7.161	7.060	7.260	800.590	750.000	6.7
Dalapon	2.600	2.500	2.700	717.500	682.500	5.1
DICAMBA	7.345	7.243	7.443	765.710	705.000	8.6
DICHLORPROP	8.044	7.942	8.142	740.860	705.000	5.1
Dinoseb	11.199	11.094	11.294	754.060	705.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/26/2025

Lab Sample No.: HSTDCCC750 Data File : PS029291.D Time Analyzed: 12:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.162	10.070	10.270	757.150	712.500	6.3
2,4,5-TP(Silvex)	9.748	9.656	9.856	757.850	712.500	6.4
2,4-D	8.852	8.761	8.961	734.420	705.000	4.2
2,4-DB	10.724	10.634	10.834	780.800	712.500	9.6
2,4-DCAA	7.624	7.532	7.732	771.960	750.000	2.9
Dalapon	2.640	2.551	2.751	675.780	682.500	-1.0
DICAMBA	7.818	7.726	7.926	745.380	705.000	5.7
DICHLORPROP	8.528	8.436	8.636	723.010	705.000	2.6
Dinoseb	11.100	11.009	11.209	737.790	705.000	4.7