

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS021925

Review By	Abdul	Review On	2/20/2025 9:47:17 AM
Supervise By	Ankita	Supervise On	2/20/2025 1:02:47 PM
SubDirectory	PS021925	HP Acquire Method	HP Processing Method ps021925 81512

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029189.D	19 Feb 2025 09:03		AR\AJ	Ok
2	I.BLK	I.BLK	PS029190.D	19 Feb 2025 09:27		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029191.D	19 Feb 2025 10:21		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029192.D	19 Feb 2025 10:45		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029193.D	19 Feb 2025 11:09		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS029194.D	19 Feb 2025 11:33		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS029195.D	19 Feb 2025 11:57		AR\AJ	Ok
8	HSTDICV750	ICVPS021925	PS029196.D	19 Feb 2025 12:21		AR\AJ	Ok
9	I.BLK	I.BLK	PS029197.D	19 Feb 2025 12:45		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029198.D	19 Feb 2025 13:09		AR\AJ	Ok
11	PB166753BL	PB166753BL	PS029199.D	19 Feb 2025 13:33		AR\AJ	Ok
12	PB166753BS	PB166753BS	PS029200.D	19 Feb 2025 13:57		AR\AJ	Ok
13	Q1373-01	VB27161	PS029201.D	19 Feb 2025 14:21		AR\AJ	Ok,M
14	Q1373-01MS	VB27161MS	PS029202.D	19 Feb 2025 14:45	Comp#1,14 recovery fail	AR\AJ	Ok,M
15	Q1373-01MSD	VB27161MSD	PS029203.D	19 Feb 2025 15:09	Comp#1,14 recovery fail	AR\AJ	Ok,M
16	Q1373-02	VDR25992	PS029204.D	19 Feb 2025 15:33		AR\AJ	Ok,M
17	PB166735TB	PB166735TB	PS029205.D	19 Feb 2025 15:57		AR\AJ	Ok,M
18	I.BLK	I.BLK	PS029206.D	19 Feb 2025 16:21		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS021925

Review By	Abdul	Review On	2/20/2025 9:47:17 AM
Supervise By	Ankita	Supervise On	2/20/2025 1:02:47 PM
SubDirectory	PS021925	HP Acquire Method	HP Processing Method ps021925 81512
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC	PP24066		
Internal Standard/PEM	PP24069,PP24070		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	HSTDCCC750	HSTDCCC750	PS029207.D	19 Feb 2025 16:45		AR\AJ	Ok
20	PB166763BL	PB166763BL	PS029208.D	19 Feb 2025 17:09		AR\AJ	Ok
21	PB166763BS	PB166763BS	PS029209.D	19 Feb 2025 17:33		AR\AJ	Ok
22	Q1372-01	AU-05-021425	PS029210.D	19 Feb 2025 17:57		AR\AJ	Ok,M
23	Q1372-01MS	AU-05-021425MS	PS029211.D	19 Feb 2025 18:21	some compound recovery fail	AR\AJ	Ok,M
24	Q1372-01MSD	AU-05-021425MSD	PS029212.D	19 Feb 2025 18:45	some compound recovery fail , RPD fail	AR\AJ	Ok,M
25	Q1382-01	OU4-PCS-TC-01-02172	PS029213.D	19 Feb 2025 19:09		AR\AJ	Ok,M
26	Q1382-03	OU4-PCS-TC-02-02172	PS029214.D	19 Feb 2025 19:33		AR\AJ	Ok
27	Q1382-05	OU4-PCS-TC-03-02172	PS029215.D	19 Feb 2025 19:57		AR\AJ	Ok,M
28	Q1382-07	OU4-PCS-TC-04-02172	PS029216.D	19 Feb 2025 20:21		AR\AJ	Ok
29	Q1382-09	OU4-PCS-TC-05-02172	PS029217.D	19 Feb 2025 20:45		AR\AJ	Ok
30	I.BLK	I.BLK	PS029218.D	19 Feb 2025 21:09		AR\AJ	Ok
31	HSTDCCC750	HSTDCCC750	PS029219.D	19 Feb 2025 21:33		AR\AJ	Ok,M
32	Q1382-11	OU4-PCS-TC-06-02172	PS029220.D	19 Feb 2025 22:21	END CCC Fail	AR\AJ	Not Ok
33	Q1382-13	OU4-PCS-TC-07-02172	PS029221.D	19 Feb 2025 22:45	END CCC Fail	AR\AJ	Not Ok
34	Q1382-15	OU4-PCS-TC-08-02172	PS029222.D	19 Feb 2025 23:09	END CCC Fail	AR\AJ	Not Ok
35	Q1382-17	OU4-PCS-TC-09-02172	PS029223.D	19 Feb 2025 23:33	END CCC Fail	AR\AJ	Not Ok
36	Q1382-19	OU4-PCS-TC-10-02172	PS029224.D	19 Feb 2025 23:57	END CCC Fail	AR\AJ	Not Ok
37	PB166764BL	PB166764BL	PS029225.D	20 Feb 2025 00:21	END CCC Fail	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS021925

Review By	Abdul	Review On	2/20/2025 9:47:17 AM
Supervise By	Ankita	Supervise On	2/20/2025 1:02:47 PM
SubDirectory	PS021925	HP Acquire Method	HP Processing Method ps021925 81512
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC	PP24066		
Internal Standard/PEM	PP24069,PP24070		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB166764BS	PB166764BS	PS029226.D	20 Feb 2025 00:45	END CCC Fail	AR\AJ	Not Ok
39	Q1383-01	OU4-PCS-TC-11-02172	PS029227.D	20 Feb 2025 01:09	END CCC Fail	AR\AJ	Not Ok
40	Q1383-01MS	OU4-PCS-TC-11-02172	PS029228.D	20 Feb 2025 01:33	END CCC Fail , Some comp recovery fail	AR\AJ	Not Ok
41	Q1383-01MSD	OU4-PCS-TC-11-02172	PS029229.D	20 Feb 2025 01:56	END CCC Fail , Some comp recovery fail	AR\AJ	Not Ok
42	I.BLK	I.BLK	PS029230.D	20 Feb 2025 02:20		AR\AJ	Not Ok
43	HSTDCCC750	HSTDCCC750	PS029231.D	20 Feb 2025 02:44	CCC Fail	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS022125

Review By	Abdul	Review On	2/22/2025 8:32:12 PM
Supervise By	Ankita	Supervise On	2/24/2025 3:48:02 PM
SubDirectory	PS022125	HP Acquire Method	HP Processing Method ps0221258151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029232.D	21 Feb 2025 19:08		AR\AJ	Ok
2	I.BLK	I.BLK	PS029233.D	21 Feb 2025 19:32		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029234.D	21 Feb 2025 19:56		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029235.D	21 Feb 2025 20:20		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029236.D	21 Feb 2025 20:44	Method fail for comp 10	AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS029237.D	21 Feb 2025 21:08		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS029238.D	21 Feb 2025 21:32		AR\AJ	Ok
8	HSTDICV750	ICVPS022125	PS029239.D	21 Feb 2025 21:56		AR\AJ	Ok
9	I.BLK	I.BLK	PS029240.D	21 Feb 2025 22:20		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029241.D	21 Feb 2025 22:44		AR\AJ	Ok
11	Q1383-03	OU4-PCS-TC-12-02172	PS029242.D	21 Feb 2025 23:08		AR\AJ	Ok,M
12	Q1383-05	OU4-PCS-TC-13-02172	PS029243.D	21 Feb 2025 23:32		AR\AJ	Ok,M
13	Q1383-07	OU4-PCS-TC-14-02172	PS029244.D	21 Feb 2025 23:56		AR\AJ	Ok,M
14	Q1383-09	OU4-PCS-TC-15-02172	PS029245.D	22 Feb 2025 00:20		AR\AJ	Ok,M
15	Q1383-11	OU4-PCS-TC-16-02172	PS029246.D	22 Feb 2025 00:44		AR\AJ	Ok,M
16	Q1383-13	OU4-PCS-TC-17-02172	PS029247.D	22 Feb 2025 01:08	2,4-DCAA high in 1st column	AR\AJ	ReRun
17	Q1383-15	OU4-PCS-TC-18-02172	PS029248.D	22 Feb 2025 01:32		AR\AJ	Ok,M
18	Q1383-17	OU4-PCS-TC-19-02172	PS029249.D	22 Feb 2025 01:56		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS022125

Review By	Abdul	Review On	2/22/2025 8:32:12 PM
Supervise By	Ankita	Supervise On	2/24/2025 3:48:02 PM
SubDirectory	PS022125	HP Acquire Method	HP Processing Method ps0221258151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

19	Q1383-19	OU4-PCS-TC-20-02172	PS029250.D	22 Feb 2025 02:20		AR\AJ	Ok
20	I.BLK	I.BLK	PS029251.D	22 Feb 2025 02:44		AR\AJ	Ok
21	HSTDCCC750	HSTDCCC750	PS029252.D	22 Feb 2025 03:08		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS022625

Review By	Abdul	Review On	2/27/2025 5:22:52 PM
Supervise By	Ankita	Supervise On	2/28/2025 10:01:46 AM
SubDirectory	PS022625	HP Acquire Method	HP Processing Method ps0221258151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029289.D	26 Feb 2025 09:29		AR\AJ	Ok
2	I.BLK	I.BLK	PS029290.D	26 Feb 2025 09:53		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS029291.D	26 Feb 2025 12:55		AR\AJ	Ok,M
4	Q1383-21	OU4-CF-15-021725	PS029292.D	26 Feb 2025 14:58		AR\AJ	Ok,M
5	Q1383-01	OU4-PCS-TC-11-021725	PS029293.D	26 Feb 2025 15:22		AR\AJ	Ok,M
6	Q1383-01MS	OU4-PCS-TC-11-021725	PS029294.D	26 Feb 2025 15:46	some compound recovery fail	AR\AJ	Ok,M
7	Q1383-01MSD	OU4-PCS-TC-11-021725	PS029295.D	26 Feb 2025 16:10	some compound recovery fail	AR\AJ	Ok,M
8	Q1382-11	OU4-PCS-TC-06-021725	PS029296.D	26 Feb 2025 16:34		AR\AJ	Ok,M
9	Q1382-13	OU4-PCS-TC-07-021725	PS029297.D	26 Feb 2025 16:58		AR\AJ	Ok,M
10	Q1382-15	OU4-PCS-TC-08-021725	PS029298.D	26 Feb 2025 17:22		AR\AJ	Ok,M
11	Q1382-17	OU4-PCS-TC-09-021725	PS029299.D	26 Feb 2025 17:46		AR\AJ	Ok,M
12	Q1382-19	OU4-PCS-TC-10-021725	PS029300.D	26 Feb 2025 18:10		AR\AJ	Ok,M
13	PB166764BL	PB166764BL	PS029301.D	26 Feb 2025 18:34		AR\AJ	Ok,M
14	I.BLK	I.BLK	PS029302.D	26 Feb 2025 18:59		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS029303.D	26 Feb 2025 19:23		AR\AJ	Ok,M
16	PB166764BS	PB166764BS	PS029304.D	26 Feb 2025 20:35		AR\AJ	Ok
17	I.BLK	I.BLK	PS029305.D	26 Feb 2025 20:59		AR\AJ	Ok
18	HSTDCCC750	HSTDCCC750	PS029306.D	26 Feb 2025 21:23		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS022625

Review By	Abdul	Review On	2/27/2025 5:22:52 PM
Supervise By	Ankita	Supervise On	2/28/2025 10:01:46 AM
SubDirectory	PS022625	HP Acquire Method	HP Processing Method ps0221258151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

M : Manual Integration