

Surrogate Summary

SDG No.: Q1382

Client: Nobis Group

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029190.D	PIBLK-PS029190.D	2,4-DCAA	1	500	492	98		32	138
		2,4-DCAA	2	500	499	100		32	138
I.BLK-PS029206.D	PIBLK-PS029206.D	2,4-DCAA	1	500	487	97		32	138
		2,4-DCAA	2	500	507	101		32	138
PB166763BL	PB166763BL	2,4-DCAA	1	500	466	93		27	122
		2,4-DCAA	2	500	475	95		27	122
PB166763BS	PB166763BS	2,4-DCAA	1	500	516	103		27	122
		2,4-DCAA	2	500	523	105		27	122
Q1372-01MS	AU-05-021425MS	2,4-DCAA	1	500	203	41		27	122
		2,4-DCAA	2	500	144	29		27	122
Q1372-01MSD	AU-05-021425MSD	2,4-DCAA	1	500	202	40		27	122
		2,4-DCAA	2	500	136	27		27	122
Q1382-01	OU4-PCS-TC-01-021725	2,4-DCAA	1	500	417	83		27	122
		2,4-DCAA	2	500	361	72		27	122
Q1382-03	OU4-PCS-TC-02-021725	2,4-DCAA	1	500	434	87		27	122
		2,4-DCAA	2	500	403	81		27	122
Q1382-05	OU4-PCS-TC-03-021725	2,4-DCAA	1	500	433	87		27	122
		2,4-DCAA	2	500	383	77		27	122
Q1382-07	OU4-PCS-TC-04-021725	2,4-DCAA	1	500	380	76		27	122
		2,4-DCAA	2	500	366	73		27	122
Q1382-09	OU4-PCS-TC-05-021725	2,4-DCAA	1	500	420	84		27	122
		2,4-DCAA	2	500	416	83		27	122
I.BLK-PS029218.D	PIBLK-PS029218.D	2,4-DCAA	1	500	438	88		32	138
		2,4-DCAA	2	500	500	100		32	138
I.BLK-PS029233.D	PIBLK-PS029233.D	2,4-DCAA	1	500	476	95		32	138
		2,4-DCAA	2	500	494	99		32	138
I.BLK-PS029290.D	PIBLK-PS029290.D	2,4-DCAA	1	500	522	104		32	138
		2,4-DCAA	2	500	511	102		32	138
Q1382-11	OU4-PCS-TC-06-021725	2,4-DCAA	1	500	443	89		27	122
		2,4-DCAA	2	500	453	91		27	122
Q1382-13	OU4-PCS-TC-07-021725	2,4-DCAA	1	500	521	104		27	122
		2,4-DCAA	2	500	553	111		27	122
Q1382-15	OU4-PCS-TC-08-021725	2,4-DCAA	1	500	550	110		27	122
		2,4-DCAA	2	500	562	112		27	122
Q1382-17	OU4-PCS-TC-09-021725	2,4-DCAA	1	500	539	108		27	122
		2,4-DCAA	2	500	528	106		27	122
Q1382-19	OU4-PCS-TC-10-021725	2,4-DCAA	1	500	502	100		27	122
		2,4-DCAA	2	500	476	95		27	122
I.BLK-PS029302.D	PIBLK-PS029302.D	2,4-DCAA	1	500	531	106		32	138
		2,4-DCAA	2	500	527	105		32	138