

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 02:15 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.67       | 5.69      | 5.59      | 5.79  | 0.02       |
| Aroclor-1248-2 (2)   | 5.95       | 5.97      | 5.87      | 6.07  | 0.02       |
| Aroclor-1248-3 (3)   | 6.15       | 6.17      | 6.07      | 6.27  | 0.02       |
| Aroclor-1248-4 (4)   | 6.55       | 6.57      | 6.47      | 6.67  | 0.02       |
| Aroclor-1248-5 (5)   | 6.59       | 6.61      | 6.51      | 6.71  | 0.02       |
| Tetrachloro-m-xylene | 4.52       | 4.54      | 4.44      | 4.64  | 0.02       |
| Decachlorobiphenyl   | 10.25      | 10.28     | 10.18     | 10.38 | 0.03       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 02:15 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.92       | 4.93      | 4.83      | 5.03 | 0.01       |
| Aroclor-1248-2       | (2) | 5.16       | 5.17      | 5.07      | 5.27 | 0.01       |
| Aroclor-1248-3       | (3) | 5.20       | 5.22      | 5.12      | 5.32 | 0.02       |
| Aroclor-1248-4       | (4) | 5.37       | 5.39      | 5.29      | 5.49 | 0.02       |
| Aroclor-1248-5       | (5) | 5.77       | 5.78      | 5.68      | 5.88 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.84      | 3.74      | 3.94 | 0.01       |
| Decachlorobiphenyl   |     | 8.89       | 8.91      | 8.81      | 9.01 | 0.02       |