

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 07:07 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1254-1       | (1) | 6.53       | 6.54      | 6.44                   | 6.64  | 0.01       |
| Aroclor-1254-2       | (2) | 6.75       | 6.76      | 6.66                   | 6.86  | 0.02       |
| Aroclor-1254-3       | (3) | 7.11       | 7.12      | 7.02                   | 7.22  | 0.01       |
| Aroclor-1254-4       | (4) | 7.39       | 7.40      | 7.30                   | 7.50  | 0.01       |
| Aroclor-1254-5       | (5) | 7.81       | 7.82      | 7.72                   | 7.92  | 0.01       |
| Tetrachloro-m-xylene |     | 4.53       | 4.53      | 4.43                   | 4.63  | 0.00       |
| Decachlorobiphenyl   |     | 10.26      | 10.28     | 10.18                  | 10.38 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 07:07 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.73       | 5.74      | 5.64      | 5.84 | 0.01       |
| Aroclor-1254-2       | (2) | 5.88       | 5.89      | 5.79      | 5.99 | 0.01       |
| Aroclor-1254-3       | (3) | 6.28       | 6.30      | 6.20      | 6.40 | 0.02       |
| Aroclor-1254-4       | (4) | 6.51       | 6.52      | 6.42      | 6.62 | 0.01       |
| Aroclor-1254-5       | (5) | 6.93       | 6.94      | 6.84      | 7.04 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.84      | 3.74      | 3.94 | 0.01       |
| Decachlorobiphenyl   |     | 8.89       | 8.91      | 8.81      | 9.01 | 0.02       |