

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 17:52 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.53	6.54	6.44	6.64	0.01
Aroclor-1254-2	(2)	6.75	6.76	6.66	6.86	0.01
Aroclor-1254-3	(3)	7.11	7.12	7.02	7.22	0.01
Aroclor-1254-4	(4)	7.39	7.40	7.30	7.50	0.01
Aroclor-1254-5	(5)	7.81	7.82	7.72	7.92	0.01
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.26	10.28	10.18	10.38	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 17:52 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.73	5.74	5.64	5.84	0.01
Aroclor-1254-2	(2)	5.88	5.89	5.79	5.99	0.01
Aroclor-1254-3	(3)	6.28	6.30	6.20	6.40	0.02
Aroclor-1254-4	(4)	6.51	6.52	6.42	6.62	0.01
Aroclor-1254-5	(5)	6.93	6.94	6.84	7.04	0.01
Tetrachloro-m-xylene		3.83	3.84	3.74	3.94	0.01
Decachlorobiphenyl		8.89	8.91	8.81	9.01	0.02