

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069830.D Time Analyzed: 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.524	6.439	6.639	505.170	500.000	1.0
Aroclor-1254-2	6.741	6.656	6.856	556.920	500.000	11.4
Aroclor-1254-3	7.104	7.019	7.219	521.840	500.000	4.4
Aroclor-1254-4	7.386	7.301	7.501	525.230	500.000	5.0
Aroclor-1254-5	7.802	7.719	7.919	518.390	500.000	3.7
Decachlorobiphenyl	10.249	10.175	10.375	52.000	50.000	4.0
Tetrachloro-m-xylene	4.522	4.434	4.634	55.430	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069830.D Time Analyzed: 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.726	5.642	5.842	520.360	500.000	4.1
Aroclor-1254-2	5.874	5.790	5.990	515.180	500.000	3.0
Aroclor-1254-3	6.279	6.195	6.395	518.900	500.000	3.8
Aroclor-1254-4	6.507	6.423	6.623	522.360	500.000	4.5
Aroclor-1254-5	6.926	6.844	7.044	522.720	500.000	4.5
Decachlorobiphenyl	8.886	8.809	9.009	48.350	50.000	-3.3
Tetrachloro-m-xylene	3.828	3.740	3.940	54.330	50.000	8.7