

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP069842.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.677	5.592	5.792	576.500	500.000	15.3
Aroclor-1016-2	5.700	5.614	5.814	540.700	500.000	8.1
Aroclor-1016-3	5.762	5.676	5.876	522.670	500.000	4.5
Aroclor-1016-4	5.859	5.774	5.974	529.490	500.000	5.9
Aroclor-1016-5	6.153	6.067	6.267	491.810	500.000	-1.6
Aroclor-1260-1	7.272	7.187	7.387	547.050	500.000	9.4
Aroclor-1260-2	7.526	7.441	7.641	507.620	500.000	1.5
Aroclor-1260-3	7.884	7.800	8.000	516.930	500.000	3.4
Aroclor-1260-4	8.109	8.025	8.225	487.880	500.000	-2.4
Aroclor-1260-5	8.429	8.347	8.547	509.920	500.000	2.0
Decachlorobiphenyl	10.251	10.179	10.379	53.100	50.000	6.2
Tetrachloro-m-xylene	4.524	4.438	4.638	55.760	50.000	11.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP069842.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.834	5.034	487.250	500.000	-2.6
Aroclor-1016-2	4.939	4.853	5.053	506.240	500.000	1.2
Aroclor-1016-3	5.117	5.030	5.230	520.360	500.000	4.1
Aroclor-1016-4	5.159	5.072	5.272	488.050	500.000	-2.4
Aroclor-1016-5	5.373	5.288	5.488	511.960	500.000	2.4
Aroclor-1260-1	6.412	6.326	6.526	557.250	500.000	11.5
Aroclor-1260-2	6.600	6.515	6.715	564.770	500.000	13.0
Aroclor-1260-3	6.754	6.670	6.870	488.800	500.000	-2.2
Aroclor-1260-4	7.226	7.142	7.342	543.010	500.000	8.6
Aroclor-1260-5	7.467	7.384	7.584	550.080	500.000	10.0
Decachlorobiphenyl	8.888	8.811	9.011	50.800	50.000	1.6
Tetrachloro-m-xylene	3.830	3.740	3.940	52.910	50.000	5.8