

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069845.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1254-1	6.530	6.439	6.639	508.870	500.000	1.8
Aroclor-1254-2	6.746	6.656	6.856	561.710	500.000	12.3
Aroclor-1254-3	7.109	7.019	7.219	525.790	500.000	5.2
Aroclor-1254-4	7.391	7.301	7.501	536.840	500.000	7.4
Aroclor-1254-5	7.808	7.719	7.919	518.660	500.000	3.7
Decachlorobiphenyl	10.253	10.175	10.375	52.330	50.000	4.7
Tetrachloro-m-xylene	4.526	4.434	4.634	55.590	50.000	11.2

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069845.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.728	5.642	5.842	535.410	500.000	7.1
Aroclor-1254-2	5.876	5.790	5.990	537.250	500.000	7.5
Aroclor-1254-3	6.281	6.195	6.395	548.920	500.000	9.8
Aroclor-1254-4	6.509	6.423	6.623	545.020	500.000	9.0
Aroclor-1254-5	6.928	6.844	7.044	549.610	500.000	9.9
Decachlorobiphenyl	8.889	8.809	9.009	49.690	50.000	-0.6
Tetrachloro-m-xylene	3.830	3.740	3.940	54.430	50.000	8.9