

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL05 Date Analyzed: 02/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP069857.D Time Analyzed: 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.592	5.792	482.440	500.000	-3.5
Aroclor-1016-2	5.702	5.614	5.814	505.860	500.000	1.2
Aroclor-1016-3	5.764	5.676	5.876	500.320	500.000	0.1
Aroclor-1016-4	5.862	5.774	5.974	517.400	500.000	3.5
Aroclor-1016-5	6.155	6.067	6.267	467.500	500.000	-6.5
Aroclor-1260-1	7.274	7.187	7.387	514.180	500.000	2.8
Aroclor-1260-2	7.528	7.441	7.641	475.840	500.000	-4.8
Aroclor-1260-3	7.886	7.800	8.000	488.150	500.000	-2.4
Aroclor-1260-4	8.111	8.025	8.225	467.550	500.000	-6.5
Aroclor-1260-5	8.431	8.347	8.547	481.930	500.000	-3.6
Decachlorobiphenyl	10.254	10.179	10.379	50.340	50.000	0.7
Tetrachloro-m-xylene	4.528	4.438	4.638	51.890	50.000	3.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL05 Date Analyzed: 02/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP069857.D Time Analyzed: 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.834	5.034	446.890	500.000	-10.6
Aroclor-1016-2	4.939	4.853	5.053	464.700	500.000	-7.1
Aroclor-1016-3	5.116	5.030	5.230	478.460	500.000	-4.3
Aroclor-1016-4	5.158	5.072	5.272	446.550	500.000	-10.7
Aroclor-1016-5	5.373	5.288	5.488	527.390	500.000	5.5
Aroclor-1260-1	6.411	6.326	6.526	500.170	500.000	0.0
Aroclor-1260-2	6.599	6.515	6.715	496.810	500.000	-0.6
Aroclor-1260-3	6.753	6.670	6.870	439.430	500.000	-12.1
Aroclor-1260-4	7.226	7.142	7.342	475.050	500.000	-5.0
Aroclor-1260-5	7.466	7.384	7.584	485.160	500.000	-3.0
Decachlorobiphenyl	8.889	8.811	9.011	44.480	50.000	-11.0
Tetrachloro-m-xylene	3.829	3.740	3.940	50.120	50.000	0.2