

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL07 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069917.D Time Analyzed: 17:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.677	5.592	5.792	498.870	500.000	-0.2
Aroclor-1016-2	5.700	5.614	5.814	524.890	500.000	5.0
Aroclor-1016-3	5.762	5.676	5.876	506.260	500.000	1.3
Aroclor-1016-4	5.860	5.774	5.974	511.780	500.000	2.4
Aroclor-1016-5	6.151	6.067	6.267	515.140	500.000	3.0
Aroclor-1260-1	7.272	7.187	7.387	528.180	500.000	5.6
Aroclor-1260-2	7.526	7.441	7.641	498.650	500.000	-0.3
Aroclor-1260-3	7.885	7.800	8.000	513.520	500.000	2.7
Aroclor-1260-4	8.109	8.025	8.225	486.170	500.000	-2.8
Aroclor-1260-5	8.429	8.347	8.547	502.940	500.000	0.6
Decachlorobiphenyl	10.252	10.179	10.379	52.360	50.000	4.7
Tetrachloro-m-xylene	4.524	4.438	4.638	53.840	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL07 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069917.D Time Analyzed: 17:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.834	5.034	507.320	500.000	1.5
Aroclor-1016-2	4.939	4.853	5.053	487.140	500.000	-2.6
Aroclor-1016-3	5.117	5.030	5.230	504.060	500.000	0.8
Aroclor-1016-4	5.158	5.072	5.272	470.460	500.000	-5.9
Aroclor-1016-5	5.373	5.288	5.488	500.290	500.000	0.1
Aroclor-1260-1	6.412	6.326	6.526	539.520	500.000	7.9
Aroclor-1260-2	6.600	6.515	6.715	548.700	500.000	9.7
Aroclor-1260-3	6.754	6.670	6.870	478.090	500.000	-4.4
Aroclor-1260-4	7.227	7.142	7.342	534.290	500.000	6.9
Aroclor-1260-5	7.467	7.384	7.584	545.220	500.000	9.0
Decachlorobiphenyl	8.889	8.811	9.011	50.550	50.000	1.1
Tetrachloro-m-xylene	3.830	3.740	3.940	52.130	50.000	4.3