

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL08 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069920.D Time Analyzed: 17:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1254-1	6.530	6.439	6.639	498.950	500.000	-0.2
Aroclor-1254-2	6.746	6.656	6.856	529.500	500.000	5.9
Aroclor-1254-3	7.109	7.019	7.219	510.030	500.000	2.0
Aroclor-1254-4	7.391	7.301	7.501	516.180	500.000	3.2
Aroclor-1254-5	7.808	7.719	7.919	509.500	500.000	1.9
Decachlorobiphenyl	10.256	10.175	10.375	52.250	50.000	4.5
Tetrachloro-m-xylene	4.527	4.434	4.634	54.740	50.000	9.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL08 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069920.D Time Analyzed: 17:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.726	5.642	5.842	487.080	500.000	-2.6
Aroclor-1254-2	5.875	5.790	5.990	492.350	500.000	-1.5
Aroclor-1254-3	6.280	6.195	6.395	499.690	500.000	-0.1
Aroclor-1254-4	6.509	6.423	6.623	537.170	500.000	7.4
Aroclor-1254-5	6.928	6.844	7.044	514.790	500.000	3.0
Decachlorobiphenyl	8.888	8.809	9.009	47.810	50.000	-4.4
Tetrachloro-m-xylene	3.829	3.740	3.940	53.290	50.000	6.6