

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL10 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069935.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.525	6.439	6.639	467.990	500.000	-6.4
Aroclor-1254-2	6.743	6.656	6.856	537.130	500.000	7.4
Aroclor-1254-3	7.106	7.019	7.219	518.010	500.000	3.6
Aroclor-1254-4	7.389	7.301	7.501	522.070	500.000	4.4
Aroclor-1254-5	7.806	7.719	7.919	515.750	500.000	3.2
Decachlorobiphenyl	10.253	10.175	10.375	52.470	50.000	4.9
Tetrachloro-m-xylene	4.524	4.434	4.634	54.940	50.000	9.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL10 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069935.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.727	5.642	5.842	496.840	500.000	-0.6
Aroclor-1254-2	5.875	5.790	5.990	504.230	500.000	0.8
Aroclor-1254-3	6.281	6.195	6.395	507.220	500.000	1.4
Aroclor-1254-4	6.509	6.423	6.623	535.150	500.000	7.0
Aroclor-1254-5	6.928	6.844	7.044	520.410	500.000	4.1
Decachlorobiphenyl	8.889	8.809	9.009	50.140	50.000	0.3
Tetrachloro-m-xylene	3.830	3.740	3.940	54.030	50.000	8.1