

Surrogate Summary

SDG No.: Q1382

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069264.D	PIBLK-PP069264.D	Tetrachloro-m-xylene	1	20	25.7	128		60	140
		Decachlorobiphenyl	1	20	25.4	127		60	140
		Tetrachloro-m-xylene	2	20	26.0	130		60	140
		Decachlorobiphenyl	2	20	25.4	127		60	140
I.BLK-PP069831.D	PIBLK-PP069831.D	Tetrachloro-m-xylene	1	20	23.3	116		60	140
		Decachlorobiphenyl	1	20	22.6	113		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
PB166770BL	PB166770BL	Tetrachloro-m-xylene	1	20	25.6	128		44	130
		Decachlorobiphenyl	1	20	22.4	112		60	125
		Tetrachloro-m-xylene	2	20	25.7	129		44	130
		Decachlorobiphenyl	2	20	23.8	119		60	125
PB166770BS	PB166770BS	Tetrachloro-m-xylene	1	20	25.5	128		44	130
		Decachlorobiphenyl	1	20	23.5	117		60	125
		Tetrachloro-m-xylene	2	20	24.3	121		44	130
		Decachlorobiphenyl	2	20	24.6	123		60	125
Q1382-01	OU4-PCS-TC-01-021725	Tetrachloro-m-xylene	1	20	25.6	128		44	130
		Decachlorobiphenyl	1	20	18.4	92		60	125
		Tetrachloro-m-xylene	2	20	25.3	127		44	130
		Decachlorobiphenyl	2	20	19.1	96		60	125
I.BLK-PP069846.D	PIBLK-PP069846.D	Tetrachloro-m-xylene	1	20	23.1	116		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	22.9	115		60	140
		Decachlorobiphenyl	2	20	22.9	114		60	140
Q1382-03	OU4-PCS-TC-02-021725	Tetrachloro-m-xylene	1	20	26.8	134	*	44	130
		Decachlorobiphenyl	1	20	23.1	115		60	125
		Tetrachloro-m-xylene	2	20	27.0	135	*	44	130
		Decachlorobiphenyl	2	20	23.2	116		60	125
Q1382-05	OU4-PCS-TC-03-021725	Tetrachloro-m-xylene	1	20	26.6	133	*	44	130
		Decachlorobiphenyl	1	20	22.7	114		60	125
		Tetrachloro-m-xylene	2	20	26.1	131	*	44	130
		Decachlorobiphenyl	2	20	23.4	117		60	125
Q1382-07	OU4-PCS-TC-04-021725	Tetrachloro-m-xylene	1	20	27.7	139	*	44	130
		Decachlorobiphenyl	1	20	23.2	116		60	125
		Tetrachloro-m-xylene	2	20	27.9	139	*	44	130
		Decachlorobiphenyl	2	20	23.4	117		60	125
Q1382-11	OU4-PCS-TC-06-021725	Tetrachloro-m-xylene	1	20	27.1	135	*	44	130
		Decachlorobiphenyl	1	20	23.1	115		60	125
		Tetrachloro-m-xylene	2	20	27.5	138	*	44	130
		Decachlorobiphenyl	2	20	22.1	111		60	125
Q1382-13	OU4-PCS-TC-07-021725	Tetrachloro-m-xylene	1	20	26.8	134	*	44	130

Surrogate Summary

SDG No.: **Q1382**

Client: **Nobis Group**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1382-13	OU4-PCS-TC-07-021725	Decachlorobiphenyl	1	20	23.4	117		60	125
		Tetrachloro-m-xylene	2	20	24.9	125		44	130
Q1382-15	OU4-PCS-TC-08-021725	Decachlorobiphenyl	2	20	23.2	116		60	125
		Tetrachloro-m-xylene	1	20	25.3	126		44	130
		Decachlorobiphenyl	1	20	19.8	99		60	125
		Tetrachloro-m-xylene	2	20	24.6	123		44	130
Q1382-17	OU4-PCS-TC-09-021725	Decachlorobiphenyl	2	20	19.7	98		60	125
		Tetrachloro-m-xylene	1	20	24.3	121		44	130
		Decachlorobiphenyl	1	20	20.0	100		60	125
		Tetrachloro-m-xylene	2	20	24.4	122		44	130
Q1382-19	OU4-PCS-TC-10-021725	Decachlorobiphenyl	2	20	20.4	102		60	125
		Tetrachloro-m-xylene	1	20	27.1	136	*	44	130
		Decachlorobiphenyl	1	20	23.7	118		60	125
		Tetrachloro-m-xylene	2	20	27.0	135	*	44	130
I.BLK-PP069861.D	PIBLK-PP069861.D	Decachlorobiphenyl	2	20	22.1	110		60	125
		Tetrachloro-m-xylene	1	20	23.1	115		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	22.5	112		60	140
I.BLK-PP069921.D	PIBLK-PP069921.D	Decachlorobiphenyl	2	20	22.1	111		60	140
		Tetrachloro-m-xylene	1	20	23.8	119		60	140
		Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	24.0	120		60	140
Q1382-01MS	OU4-PCS-TC-01-021725MS	Decachlorobiphenyl	2	20	23.7	119		60	140
		Tetrachloro-m-xylene	1	20	23.5	117		44	130
		Decachlorobiphenyl	1	20	19.7	99		60	125
		Tetrachloro-m-xylene	2	20	22.7	114		44	130
Q1382-01MSD	OU4-PCS-TC-01-021725MSD	Decachlorobiphenyl	2	20	20.6	103		60	125
		Tetrachloro-m-xylene	1	20	24.4	122		44	130
		Decachlorobiphenyl	1	20	20.1	101		60	125
		Tetrachloro-m-xylene	2	20	23.4	117		44	130
Q1382-03RE	OU4-PCS-TC-02-021725RE	Decachlorobiphenyl	2	20	20.1	101		60	125
		Tetrachloro-m-xylene	1	20	27.2	136	*	44	130
		Decachlorobiphenyl	1	20	23.6	118		60	125
		Tetrachloro-m-xylene	2	20	26.9	135	*	44	130
Q1382-05RE	OU4-PCS-TC-03-021725RE	Decachlorobiphenyl	2	20	23.6	118		60	125
		Tetrachloro-m-xylene	1	20	27.5	137	*	44	130
		Decachlorobiphenyl	1	20	23.5	117		60	125
		Tetrachloro-m-xylene	2	20	27.4	137	*	44	130
Q1382-07RE	OU4-PCS-TC-04-021725RE	Decachlorobiphenyl	2	20	22.8	114		60	125
		Tetrachloro-m-xylene	1	20	28.4	142	*	44	130
		Decachlorobiphenyl	1	20	24.0	120		60	125

Surrogate Summary

SDG No.: Q1382

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1382-07RE	OU4-PCS-TC-04-021725RE	Tetrachloro-m-xylene	2	20	27.8	139	*	44	130
		Decachlorobiphenyl	2	20	23.9	119		60	125
Q1382-11RE	OU4-PCS-TC-06-021725RE	Tetrachloro-m-xylene	1	20	28.8	144	*	44	130
		Decachlorobiphenyl	1	20	23.9	119		60	125
		Tetrachloro-m-xylene	2	20	27.9	140	*	44	130
		Decachlorobiphenyl	2	20	23.8	119		60	125
I.BLK-PP069936.D	PIBLK-PP069936.D	Tetrachloro-m-xylene	1	20	24.4	122		60	140
		Decachlorobiphenyl	1	20	24.0	120		60	140
		Tetrachloro-m-xylene	2	20	24.2	121		60	140
		Decachlorobiphenyl	2	20	23.2	116		60	140
Q1382-19RE	OU4-PCS-TC-10-021725RE	Tetrachloro-m-xylene	1	20	27.9	140	*	44	130
		Decachlorobiphenyl	1	20	24.8	124		60	125
		Tetrachloro-m-xylene	2	20	27.0	135	*	44	130
		Decachlorobiphenyl	2	20	24.3	122		60	125
Q1382-09	OU4-PCS-TC-05-021725	Tetrachloro-m-xylene	1	20	25.8	129		44	130
		Decachlorobiphenyl	1	20	23.4	117		60	125
		Tetrachloro-m-xylene	2	20	25.3	126		44	130
		Decachlorobiphenyl	2	20	22.7	114		60	125
I.BLK-PP069951.D	PIBLK-PP069951.D	Tetrachloro-m-xylene	1	20	25.6	128		60	140
		Decachlorobiphenyl	1	20	24.6	123		60	140
		Tetrachloro-m-xylene	2	20	25.5	127		60	140
		Decachlorobiphenyl	2	20	23.8	119		60	140