

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1382

Client: Nobis Group

Analytical Method: **8081B** Datafile : PL094261.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166771BS	alpha-BHC	16.65	16.7	ug/kg	100				45	137	
	beta-BHC	16.65	16.9	ug/kg	102				50	136	
	delta-BHC	16.65	16.2	ug/kg	97				47	139	
	gamma-BHC (Lindane)	16.65	16.5	ug/kg	99				49	135	
	Heptachlor	16.65	18.0	ug/kg	108				47	136	
	Aldrin	16.65	17.0	ug/kg	102				45	136	
	Heptachlor epoxide	16.65	17.8	ug/kg	107				52	136	
	Endosulfan I	16.65	18.3	ug/kg	110				53	132	
	Dieldrin	16.65	18.2	ug/kg	109				56	136	
	4,4'-DDE	16.65	18.2	ug/kg	109				56	134	
	Endrin	16.65	19.2	ug/kg	115				57	140	
	Endosulfan II	16.65	19.0	ug/kg	114				53	134	
	4,4'-DDD	16.65	18.1	ug/kg	109				56	139	
	Endosulfan sulfate	16.65	18.8	ug/kg	113				55	136	
	4,4'-DDT	16.65	20.2	ug/kg	121				50	141	
	Methoxychlor	16.65	19.2	ug/kg	115				52	143	
	Endrin ketone	16.65	18.6	ug/kg	112				55	136	
	Endrin aldehyde	16.65	17.8	ug/kg	107				35	137	
	alpha-Chlordane	16.65	18.3	ug/kg	110				54	133	
	gamma-Chlordane	16.65	18.4	ug/kg	111				53	135	