

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG No.: Q1382

Instrument ID: BNA_F Calibration Date/Time: 02/20/2025 00:05

Lab File ID: BF141693.D Init. Calib. Date(s): 02/06/2025 02/06/2025

EPA Sample No.: SSTDCCC040EC Init. Calib. Time(s): 11:07 14:14

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.276	1.393		9.2	50.0
Phenol-d6	1.612	1.669		3.5	50.0
Nitrobenzene-d5	0.383	0.384		0.3	50.0
Naphthalene	1.041	1.078		3.6	50.0
2-Methylnaphthalene	0.677	0.691		2.1	50.0
2-Fluorobiphenyl	1.298	1.372		5.7	50.0
Acenaphthylene	1.705	1.777		4.2	50.0
Acenaphthene	1.147	1.161		1.2	50.0
Fluorene	1.301	1.330		2.2	50.0
2,4,6-Tribromophenol	0.201	0.198		-1.5	50.0
Phenanthrene	1.101	1.134		3.0	50.0
Anthracene	1.107	1.131		2.2	50.0
Fluoranthene	1.167	1.058		-9.3	50.0
Pyrene	1.703	1.869		9.7	50.0
Terphenyl-d14	1.185	1.319		11.3	50.0
Benzo (a) anthracene	1.329	1.386		4.3	50.0
Chrysene	1.228	1.304		6.2	50.0
Benzo (b) fluoranthene	1.343	1.215		-9.5	50.0
Benzo (k) fluoranthene	1.147	1.183		3.1	50.0
Benzo (a) pyrene	1.087	1.111		2.2	50.0
Indeno (1,2,3-cd) pyrene	1.236	1.449		17.2	50.0
Dibenzo (a,h) anthracene	1.001	1.182		18.1	50.0
Benzo (g,h,i) perylene	1.048	1.209		15.4	50.0

All other compounds must meet a minimum RRF of 0.010.