

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG No.: Q1382
 Instrument ID: MSVOA_Y Calibration Date(s): 02/25/2025 02/25/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 12:40 16:49
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY021309.D			RRF010 = VY021310.D		RRF020 = VY021311.D			
	RRF050 = VY021312.D			RRF150 = VY021314.D		RRF100 = VY021316.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD	
Dichlorodifluoromethane	0.510	0.484	0.431	0.493	0.480	0.447	0.474	6.2	
Chloromethane	0.709	0.628	0.566	0.602	0.597	0.570	0.612	8.6	
Vinyl Chloride	0.648	0.591	0.554	0.627	0.606	0.580	0.601	5.6	
Bromomethane	0.505	0.414	0.361	0.387	0.371	0.359	0.399	13.9	
Chloroethane	0.414	0.365	0.347	0.390	0.365	0.355	0.373	6.7	
Tetrahydrofuran	0.105	0.100	0.106	0.115	0.114	0.104	0.107	5.4	
Trichlorofluoromethane	0.976	0.915	0.838	0.936	0.909	0.870	0.908	5.3	
1,1,2-Trichlorotrifluoroethane	0.546	0.529	0.491	0.537	0.541	0.508	0.525	4.1	
1,1-Dichloroethene	0.545	0.489	0.462	0.515	0.518	0.490	0.503	5.7	
Acrylonitrile	0.122	0.115	0.116	0.127	0.127	0.115	0.120	5	
Acetone	0.122	0.102	0.094	0.128	0.104	0.103	0.109	11.9	
Carbon Disulfide	1.800	1.652	1.540	1.734	1.678	1.620	1.671	5.4	
Methyl tert-butyl Ether	1.312	1.268	1.267	1.410	1.410	1.307	1.329	4.9	
Methylene Chloride	0.733	0.598	0.539	0.585	0.547	0.529	0.588	12.8	
trans-1,2-Dichloroethene	0.589	0.536	0.507	0.573	0.570	0.547	0.554	5.4	
1,1-Dichloroethane	1.101	1.032	0.961	1.083	1.058	1.007	1.040	4.9	
2-Butanone	0.162	0.155	0.156	0.183	0.171	0.159	0.164	6.5	
Carbon Tetrachloride	0.567	0.562	0.509	0.581	0.567	0.535	0.554	4.8	
2,2-Dichloropropane	1.036	0.939	0.845	0.972	0.952	0.920	0.944	6.6	
cis-1,2-Dichloroethene	0.649	0.610	0.588	0.665	0.656	0.626	0.632	4.7	
Chloroform	1.128	1.046	0.977	1.095	1.062	1.017	1.054	5.1	
1,1,1-Trichloroethane	1.020	0.956	0.889	0.986	0.979	0.927	0.960	4.8	
1,1-Dichloropropene	0.508	0.487	0.460	0.516	0.505	0.478	0.492	4.3	
Benzene	1.515	1.445	1.359	1.533	1.472	1.409	1.456	4.5	
1,2-Dichloroethane	0.422	0.417	0.393	0.434	0.419	0.394	0.413	3.9	
Trichloroethene	0.375	0.360	0.341	0.382	0.375	0.354	0.364	4.3	
1,2-Dichloropropane	0.360	0.350	0.329	0.371	0.357	0.336	0.350	4.5	
Dibromomethane	0.192	0.189	0.183	0.203	0.199	0.184	0.191	4.2	
Bromodichloromethane	0.519	0.503	0.477	0.541	0.528	0.495	0.511	4.5	
4-Methyl-2-Pentanone	0.220	0.227	0.239	0.267	0.264	0.238	0.243	7.9	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03

Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG No.: Q1382

Instrument ID: MSVOA_Y Calibration Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Calibration Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021309.D		RRF010 = VY021310.D		RRF020 = VY021311.D		RRF050 = VY021312.D		RRF150 = VY021314.D		RRF100 = VY021316.D	
COMPOUND		RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD				
Toluene		0.887	0.888	0.852	0.965	0.936	0.897	0.904	4.4				
t-1,3-Dichloropropene		0.438	0.436	0.434	0.515	0.509	0.474	0.468	8				
cis-1,3-Dichloropropene		0.541	0.526	0.528	0.591	0.581	0.546	0.552	5				
1,1,2-Trichloroethane		0.250	0.251	0.238	0.267	0.254	0.237	0.249	4.4				
1,3-Dichloropropane		0.427	0.428	0.422	0.469	0.454	0.426	0.438	4.4				
2-Hexanone		0.133	0.146	0.155	0.183	0.178	0.162	0.159	11.9				
Dibromochloromethane		0.340	0.325	0.322	0.368	0.352	0.332	0.340	5.2				
1,2-Dibromoethane		0.234	0.231	0.226	0.254	0.245	0.227	0.236	4.7				
Tetrachloroethene		0.405	0.380	0.351	0.389	0.386	0.355	0.378	5.5				
Chlorobenzene		1.171	1.111	1.039	1.176	1.159	1.078	1.122	5				
1,1,1,2-Tetrachloroethane		0.396	0.401	0.362	0.414	0.404	0.381	0.393	4.7				
Ethyl Benzene		1.938	1.941	1.826	2.130	2.115	1.979	1.988	5.8				
m/p-Xylenes		0.735	0.738	0.696	0.802	0.783	0.733	0.748	5.1				
o-Xylene		0.689	0.675	0.648	0.758	0.739	0.694	0.701	5.9				
Styrene		1.081	1.128	1.091	1.277	1.257	1.159	1.166	7.2				
Bromoform		0.221	0.221	0.211	0.246	0.236	0.214	0.225	5.9				
Isopropylbenzene		3.744	3.605	3.437	3.976	4.009	3.783	3.759	5.8				
1,1,2,2-Tetrachloroethane		0.700	0.649	0.621	0.697	0.689	0.624	0.663	5.5				
1,2,3-Trichloropropane		0.517	0.547	0.448	0.523	0.493	0.474	0.500	7.2				
Bromobenzene		0.919	0.843	0.806	0.908	0.915	0.854	0.874	5.3				
n-propylbenzene		4.532	4.403	4.216	4.830	4.842	4.546	4.561	5.3				
2-Chlorotoluene		2.741	2.558	2.412	2.715	2.730	2.551	2.618	5				
1,3,5-Trimethylbenzene		3.092	2.977	2.873	3.250	3.264	3.042	3.083	5				
4-Chlorotoluene		2.871	2.665	2.496	2.815	2.802	2.633	2.714	5.2				
tert-Butylbenzene		2.703	2.753	2.601	2.892	3.000	2.760	2.785	5.1				
1,2,4-Trimethylbenzene		3.014	2.958	2.860	3.248	3.269	3.048	3.066	5.3				
sec-Butylbenzene		4.126	4.003	3.760	4.257	4.274	3.987	4.068	4.8				
p-Isopropyltoluene		3.340	3.267	3.162	3.593	3.641	3.377	3.397	5.5				
1,3-Dichlorobenzene		1.812	1.729	1.598	1.792	1.777	1.668	1.730	4.8				
1,4-Dichlorobenzene		1.825	1.710	1.593	1.751	1.744	1.629	1.709	5				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG No.: Q1382
 Instrument ID: MSVOA_Y Calibration Date(s): 02/25/2025 02/25/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 12:40 16:49
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021309.D		RRF010 = VY021310.D		RRF020 = VY021311.D		
		RRF050 = VY021312.D		RRF150 = VY021314.D		RRF100 = VY021316.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD
n-Butylbenzene	3.078	3.050	2.931	3.421	3.481	3.218	3.196	6.8
1,2-Dichlorobenzene	1.599	1.499	1.408	1.575	1.564	1.446	1.515	5.1
1,2-Dibromo-3-Chloropropane	0.103	0.097	0.099	0.113	0.115	0.102	0.105	7.2
1,2,4-Trichlorobenzene	0.906	0.848	0.822	0.968	1.053	0.950	0.925	9.2
Hexachlorobutadiene	0.633	0.573	0.517	0.591	0.616	0.563	0.582	7
1,2,3-Trichlorobenzene	0.742	0.719	0.698	0.824	0.896	0.809	0.781	9.6
1,2-Dichloroethane-d4	0.681	0.510	0.548	0.551	0.566	0.547	0.567	10.3
Dibromofluoromethane	0.370	0.299	0.317	0.325	0.334	0.322	0.328	7.2
Toluene-d8	1.420	1.158	1.230	1.251	1.290	1.253	1.267	6.8
4-Bromofluorobenzene	0.490	0.387	0.404	0.423	0.435	0.420	0.426	8.2
trans-1,4-Dichloro-2-butene	0.210	0.209	0.208	0.244	0.255	0.225	0.225	8.9

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.