

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382
 Lab File ID: VY021236.D Date Analyzed: 02/19/2025
 Instrument ID: MSVOA_Y Time Analyzed: 14:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	200462	7.72	311138	8.63	266562	11.43
UPPER LIMIT	400924	8.219	622276	9.128	533124	11.926
LOWER LIMIT	100231	7.219	155569	8.128	133281	10.926
EPA SAMPLE NO.						
OU4-PCS-TC-01-021725	215345	7.72	410030	8.62	371656	11.42
OU4-PCS-TC-02-021725	210258	7.72	406195	8.62	371758	11.42
OU4-PCS-TC-03-021725	207032	7.71	397674	8.62	367484	11.42
VY0219SBL01	252953	7.72	478949	8.62	429832	11.43
VY0219SBS01	187422	7.71	302089	8.62	255752	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.