

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382
 Lab File ID: VY021236.D Date Analyzed: 02/19/2025
 Instrument ID: MSVOA_Y Time Analyzed: 14:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	126814	13.359				
UPPER LIMIT	253628	13.859				
LOWER LIMIT	63407	12.859				
EPA SAMPLE NO.						
OU4-PCS-TC-01-021725	143607	13.35				
OU4-PCS-TC-02-021725	144547	13.35				
OU4-PCS-TC-03-021725	147214	13.35				
VY0219SBL01	157077	13.35				
VY0219SBS01	127239	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.