

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1382 SAS No.: Q1382 SDG NO.: Q1382
 Lab File ID: VY021275.D Date Analyzed: 02/21/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	91135	13.359				
UPPER LIMIT	182270	13.859				
LOWER LIMIT	45567.5	12.859				
EPA SAMPLE NO.						
OU4-PCS-TC-04-021725	125474	13.35				
OU4-PCS-TC-05-021725	122113	13.35				
OU4-PCS-TC-06-021725	126553	13.35				
OU4-PCS-TC-07-021725	128437	13.35				
OU4-PCS-TC-08-021725	119366	13.35				
OU4-PCS-TC-09-021725	124714	13.35				
VY0221SBL01	107899	13.35				
VY0221SBS01	99790	13.36				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.