



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Nobis Group**

**Project Name: Raymark Superfund Site**

**Project # N/A**

**Chemtech Project # Q1383**

**Test Name: VOCMS Group3**

### **A. Number of Samples and Date of Receipt:**

22 Solid samples were received on 02/18/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Cyanide, Herbicide Group1, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SPLP Extraction, SPLP Mercury, SPLP MetalGroup3, SVOCMS Group3 and VOCMS Group3. This data package contains results for VOCMS Group3.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #138 68. The analysis of VOCMS Group3 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike for {VY0221SBS01} with File ID: VY021277.D met requirements for all samples except for Chloroethane[140%], Chloromethane[148%] and Vinyl chloride[141%] are failing high but no positive hit in associate samples therefore no corrective action taken.

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VY021275.D met the requirements except for Chloroethane is failing high but no positive hit in associate samples therefore no corrective action taken.

The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

**E. Additional Comments:**

As per special requirement for this project form-1 are reported in mg/kg.  
Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The SDG Q1383 is logged for VOCMS group3. Lab is not certified for Tetrahydrofuran and trans-1,4-dichloro-2-butene compounds for 8260D method.

Trip Blank was not provided with this set of samples.  
The not QT review data is reported in the Miscellaneous.  
The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_