

Matrix Spike Summary

Client:	Nobis Group	SDG No.:	Q1383
Project:	Raymark Superfund Site	Sample ID:	Q1383-21
Client ID:	OU4-CF-15-021725MS	Percent Solids for Spike Sample:	92.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.10		0.047	U	2.1	1	100		02/19/2025

Matrix Spike Summary

Client:	Nobis Group	SDG No.:	Q1383
Project:	Raymark Superfund Site	Sample ID:	Q1383-21
Client ID:	OU4-CF-15-021725MSD	Percent Solids for Spike Sample:	92.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.00		0.047	U	2.1	1	95		02/19/2025