



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group SDG No.: Q1383
Contract: NOBI03 Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	40 - 160	CV	02/21/2025	13:04	LB134771
CRI01	Aluminum	103	100	103	40 - 160	P	03/04/2025	10:22	LB134892
	Antimony	51.1	50.0	102	40 - 160	P	03/04/2025	10:22	LB134892
	Arsenic	20.7	20.0	103	40 - 160	P	03/04/2025	10:22	LB134892
	Barium	88.5	100	88	40 - 160	P	03/04/2025	10:22	LB134892
	Beryllium	6.05	6.0	101	40 - 160	P	03/04/2025	10:22	LB134892
	Cadmium	5.97	6.0	100	40 - 160	P	03/04/2025	10:22	LB134892
	Calcium	1960	2000	98	40 - 160	P	03/04/2025	10:22	LB134892
	Chromium	10.6	10.0	106	40 - 160	P	03/04/2025	10:22	LB134892
	Cobalt	29.4	30.0	98	40 - 160	P	03/04/2025	10:22	LB134892
	Copper	22.7	20.0	114	40 - 160	P	03/04/2025	10:22	LB134892
	Iron	95.3	100	95	40 - 160	P	03/04/2025	10:22	LB134892
	Lead	11.7	12.0	98	40 - 160	P	03/04/2025	10:22	LB134892
	Magnesium	2080	2000	104	40 - 160	P	03/04/2025	10:22	LB134892
	Manganese	20.5	20.0	102	40 - 160	P	03/04/2025	10:22	LB134892
	Nickel	39.7	40.0	99	40 - 160	P	03/04/2025	10:22	LB134892
	Potassium	1860	2000	93	40 - 160	P	03/04/2025	10:22	LB134892
	Selenium	18.9	20.0	95	40 - 160	P	03/04/2025	10:22	LB134892
	Silver	10.1	10.0	101	40 - 160	P	03/04/2025	10:22	LB134892
	Sodium	1870	2000	93	40 - 160	P	03/04/2025	10:22	LB134892
	Thallium	40.0	40.0	100	40 - 160	P	03/04/2025	10:22	LB134892
	Vanadium	37.8	40.0	94	40 - 160	P	03/04/2025	10:22	LB134892
	Zinc	42.5	40.0	106	40 - 160	P	03/04/2025	10:22	LB134892