



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

# Metals

- 2b -

## CRDL STANDARD FOR AA & ICP

Client: Nobis Group SDG No.: Q1383  
Contract: NOBI03 Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383  
Initial Calibration Source: \_\_\_\_\_  
Continuing Calibration Source: \_\_\_\_\_

| Sample ID | Analyte   | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|--------------------|---------------|---------------------------|----|------------------|------------------|---------------|
| CRA       | Mercury   | 0.18           | 0.2                | 88            | 40 - 160                  | CV | 02/24/2025       | 10:43            | LB134788      |
| CRI       | Antimony  | 1.95           | 2.0                | 98            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Arsenic   | 0.99           | 1.0                | 99            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Barium    | 10.2           | 10.0               | 102           | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Beryllium | 1.04           | 1.0                | 104           | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Cadmium   | 1.05           | 1.0                | 105           | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Chromium  | 1.88           | 2.0                | 94            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Copper    | 1.98           | 2.0                | 99            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Lead      | 1.03           | 1.0                | 103           | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Nickel    | 1.02           | 1.0                | 102           | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Selenium  | 4.75           | 5.0                | 95            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Silver    | 0.97           | 1.0                | 97            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Thallium  | 0.94           | 1.0                | 94            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Vanadium  | 4.74           | 5.0                | 95            | 70 - 130                  | P  | 02/24/2025       | 12:59            | LB134802      |
|           | Zinc      | 5.27           | 5.0                | 105           | 50 - 150                  | P  | 02/24/2025       | 12:59            | LB134802      |