



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1383</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1383</u>	SAS No.:	<u>Q1383</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB14	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	10:36	LB134788

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1383</u>								
Contract: <u>NOBI03</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1383</u>			SAS No.: <u>Q1383</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	10:40	LB134788
CCB08	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	11:11	LB134788
CCB09	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	11:38	LB134788
CCB10	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	12:05	LB134788
CCB11	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/24/2025	12:40	LB134788

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group SDG No.: Q1383
Contract: NOBI03 Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/24/2025	12:35	LB134802
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:35	LB134802
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/24/2025	12:35	LB134802
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:35	LB134802
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:35	LB134802
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/24/2025	12:35	LB134802
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/24/2025	12:35	LB134802
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/24/2025	12:35	LB134802
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:35	LB134802
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/24/2025	12:35	LB134802
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:35	LB134802
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:35	LB134802
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/24/2025	12:35	LB134802
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/24/2025	12:35	LB134802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1383</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1383</u>	SAS No.: <u>Q1383</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/24/2025	12:55	LB134802
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:55	LB134802
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/24/2025	12:55	LB134802
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:55	LB134802
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:55	LB134802
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/24/2025	12:55	LB134802
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/24/2025	12:55	LB134802
	Lead	0.19	+/-1.00	J	0.75	1.00	P	02/24/2025	12:55	LB134802
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	12:55	LB134802
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/24/2025	12:55	LB134802
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:55	LB134802
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	12:55	LB134802
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/24/2025	12:55	LB134802
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/24/2025	12:55	LB134802
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/24/2025	13:41	LB134802
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	13:41	LB134802
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/24/2025	13:41	LB134802
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	13:41	LB134802
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	13:41	LB134802
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/24/2025	13:41	LB134802
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/24/2025	13:41	LB134802
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/24/2025	13:41	LB134802
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	13:41	LB134802
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/24/2025	13:41	LB134802
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	13:41	LB134802
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	13:41	LB134802
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/24/2025	13:41	LB134802
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/24/2025	13:41	LB134802
CCB03	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/24/2025	14:26	LB134802
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:26	LB134802
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/24/2025	14:26	LB134802
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:26	LB134802
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:26	LB134802
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/24/2025	14:26	LB134802
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/24/2025	14:26	LB134802
	Lead	0.11	+/-1.00	J	0.75	1.00	P	02/24/2025	14:26	LB134802
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:26	LB134802
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/24/2025	14:26	LB134802
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:26	LB134802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1383</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1383</u>			SAS No.: <u>Q1383</u>					
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:26	LB134802
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/24/2025	14:26	LB134802
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/24/2025	14:26	LB134802
CCB04	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/24/2025	14:57	LB134802
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:57	LB134802
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/24/2025	14:57	LB134802
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:57	LB134802
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:57	LB134802
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/24/2025	14:57	LB134802
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/24/2025	14:57	LB134802
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/24/2025	14:57	LB134802
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/24/2025	14:57	LB134802
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/24/2025	14:57	LB134802
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:57	LB134802
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/24/2025	14:57	LB134802
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/24/2025	14:57	LB134802
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/24/2025	14:57	LB134802