

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1383

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166842BL		WATER		Batch Number:		PB166842		Prep Date:	02/24/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	02/24/2025	11:31	LB134788
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166842TB		WATER		Batch Number:		PB166842		Prep Date:	02/24/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	02/24/2025	12:19	LB134788

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q1383

Instrument: P8

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166779BL	WATER			Batch Number:		PB166779		Prep Date:	02/19/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	02/24/2025	13:04	LB134802
	Arsenic	1.00	<1.00	U	0.25	1.00	P	02/24/2025	13:04	LB134802
	Barium	10.0	<10.0	U	1.25	10.0	P	02/24/2025	13:04	LB134802
	Beryllium	1.00	<1.00	U	0.25	1.00	P	02/24/2025	13:04	LB134802
	Cadmium	1.00	<1.00	U	0.50	1.00	P	02/24/2025	13:04	LB134802
	Chromium	2.00	<2.00	U	0.75	2.00	P	02/24/2025	13:04	LB134802
	Copper	2.00	<2.00	U	1.50	2.00	P	02/24/2025	13:04	LB134802
	Lead	1.00	<1.00	U	0.75	1.00	P	02/24/2025	13:04	LB134802
	Nickel	1.00	<1.00	U	0.25	1.00	P	02/24/2025	13:04	LB134802
	Selenium	5.00	<5.00	U	4.50	5.00	P	02/24/2025	13:04	LB134802
	Silver	1.00	<1.00	U	0.50	1.00	P	02/24/2025	13:04	LB134802
	Thallium	1.00	<1.00	U	0.50	1.00	P	02/24/2025	13:04	LB134802
	Vanadium	5.00	<5.00	U	0.25	5.00	P	02/24/2025	13:04	LB134802
	Zinc	5.00	<5.00	U	1.50	5.00	P	02/24/2025	13:04	LB134802
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166779TB	WATER			Batch Number:		PB166779		Prep Date:	02/19/2025	
	Antimony	2.00	<2.00	U	0.25	2.00	P	02/24/2025	14:43	LB134802
	Arsenic	1.00	<1.00	U	0.25	1.00	P	02/24/2025	14:43	LB134802
	Barium	10.0	<10.0	U	1.25	10.0	P	02/24/2025	14:43	LB134802
	Beryllium	1.00	<1.00	U	0.25	1.00	P	02/24/2025	14:43	LB134802
	Cadmium	1.00	<1.00	U	0.50	1.00	P	02/24/2025	14:43	LB134802
	Chromium	2.00	<2.00	U	0.75	2.00	P	02/24/2025	14:43	LB134802
	Copper	2.00	<2.00	U	1.50	2.00	P	02/24/2025	14:43	LB134802
	Lead	0.12	<1.00	J	0.75	1.00	P	02/24/2025	14:43	LB134802
	Nickel	0.18	<1.00	J	0.25	1.00	P	02/24/2025	14:43	LB134802
	Selenium	5.00	<5.00	U	4.50	5.00	P	02/24/2025	14:43	LB134802
	Silver	1.00	<1.00	U	0.50	1.00	P	02/24/2025	14:43	LB134802
	Thallium	1.00	<1.00	U	0.50	1.00	P	02/24/2025	14:43	LB134802
	Vanadium	5.00	<5.00	U	0.25	5.00	P	02/24/2025	14:43	LB134802
	Zinc	0.85	<5.00	J	1.50	5.00	P	02/24/2025	14:43	LB134802