

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>Nobis Group</u>	SDG No.:	<u>Q1383</u>
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q1383</u>
		SAS No.:	<u>Q1383</u>
		Instrument ID:	<u>P8</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Antimony	1.06	1.5	71	-2.5	5.5	02/24/2025	12:40	LB134802
	Arsenic	0.29	0.1	290	-1.9	2.1	02/24/2025	12:40	LB134802
	Barium	1.40	1.2	117	-18.8	21.2	02/24/2025	12:40	LB134802
	Beryllium	0.27			-2	2	02/24/2025	12:40	LB134802
	Cadmium	0.30	0.7	43	-1.3	2.7	02/24/2025	12:40	LB134802
	Chromium	18.7	21.0	89	17	25	02/24/2025	12:40	LB134802
	Copper	7.78	8.0	97	4	12	02/24/2025	12:40	LB134802
	Lead	4.28	4.0	107	2	6	02/24/2025	12:40	LB134802
	Nickel	5.04	6.0	84	4	8	02/24/2025	12:40	LB134802
	Selenium	0	0.3		-9.7	10	02/24/2025	12:40	LB134802
	Silver	0.070			-2	2	02/24/2025	12:40	LB134802
	Thallium	0.060			-2	2	02/24/2025	12:40	LB134802
ICSAB01	Vanadium	0.15	0.5	30	-9.5	10.5	02/24/2025	12:40	LB134802
	Zinc	10.9	11.0	99	7	15	02/24/2025	12:40	LB134802
	Antimony	20.7	22.0	94	18	26	02/24/2025	12:44	LB134802
	Arsenic	20.3	19.0	107	16.2	21.9	02/24/2025	12:44	LB134802
	Barium	21.4	22.0	97	2	42	02/24/2025	12:44	LB134802
	Beryllium	20.4	19.0	107	16.2	21.9	02/24/2025	12:44	LB134802
	Cadmium	19.4	20.0	97	17	23	02/24/2025	12:44	LB134802
	Chromium	38.0	40.0	95	34	46	02/24/2025	12:44	LB134802
	Copper	27.6	25.0	110	21	29	02/24/2025	12:44	LB134802
	Lead	23.9	25.0	96	21.3	28.8	02/24/2025	12:44	LB134802
	Nickel	24.0	24.0	100	20.4	27.6	02/24/2025	12:44	LB134802
	Selenium	20.9	19.0	110	9	29	02/24/2025	12:44	LB134802
	Silver	18.3	18.0	102	15.3	20.7	02/24/2025	12:44	LB134802
	Thallium	19.9	21.0	95	17.9	24.2	02/24/2025	12:44	LB134802
	Vanadium	19.5	19.0	103	9	29	02/24/2025	12:44	LB134802
	Zinc	30.5	29.0	105	25	33	02/24/2025	12:44	LB134802