

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1383

Sas No.: Q1383 SDG No.: Q1383

Instrument ID: P8

Start Date : 02/24/2025

Run Number: LB134802

End Date : 02/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1105	100		100							
S2	S2	1112	100		100							
S3	S3	1115	100		100							
S4	S4	1118	100		98							
S5	S5	1121	100		97							
S6	S6	1124	101		97							
S7	S7	1127	101		95							
S8	S8	1130	89		78							
ICV01	ICV01	1202	97		97							
LLICV	LLICV	1231	95		95							
ICB01	ICB01	1235	95		95							
ICSA01	ICSA01	1240	95		88							
ICSAB01	ICSAB01	1244	95		89							
CCV01	CCV01	1251	92		83							
CCB01	CCB01	1255	97		95							
CRI	CRI	1259	98		97							
PB166779BL	PB166779BL	1304	96		96							
PB166779BS	PB166779BS	1308	96		91							
Q1383-02	OU4-PCS-TC-1	1315	97		94							
Q1383-04	OU4-PCS-TC-1	1318	98		95							
Q1383-06	OU4-PCS-TC-1	1321	91		89							
Q1383-08	OU4-PCS-TC-1	1324	98		95							
Q1383-10	OU4-PCS-TC-1	1328	97		94							
Q1383-12	OU4-PCS-TC-1	1331	98		94							
CCV02	CCV02	1334	93		83							
CCB02	CCB02	1341	97		97							
Q1383-16	OU4-PCS-TC-1	1355	98		94							
Q1383-18	OU4-PCS-TC-1	1358	99		96							
Q1383-20	OU4-PCS-TC-2	1401	99		94							
Q1383-14	OU4-PCS-TC-1	1405	98		94							
Q1383-22	OU4-CF-15-02	1408	97		95							
Q1383-22DUP	OU4-CF-15-02	1411	97		93							
Q1383-22L	OU4-CF-15-02	1414	97		95							
Q1383-22MS	OU4-CF-15-02	1418	96		92							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1383

Sas No.: Q1383 SDG No.: Q1383

Instrument ID: P8

Start Date : 02/24/2025

Run Number: LB134802

End Date : 02/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1421	93		83							
CCB03	CCB03	1426	95		94							
Q1383-22MSD	OU4-CF-15-02	1430	96		93							
Q1383-22A	OU4-CF-15-02	1433	96		93							
PB166779TB	PB166779TB	1443	94		93							
CCV04	CCV04	1451	88		79							
CCB04	CCB04	1457	92		91							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1383

Sas No.: Q1383 SDG No.: Q1383

Instrument ID: P8

Start Date : 02/24/2025

Run Number: LB134802

End Date : 02/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1105	100									
S2	S2	1112	99									
S3	S3	1115	97									
S4	S4	1118	96									
S5	S5	1121	94									
S6	S6	1124	94									
S7	S7	1127	91									
S8	S8	1130	80									
ICV01	ICV01	1202	96									
LLICV	LLICV	1231	93									
ICB01	ICB01	1235	94									
ICSA01	ICSA01	1240	87									
ICSAB01	ICSAB01	1244	88									
CCV01	CCV01	1251	84									
CCB01	CCB01	1255	96									
CRI	CRI	1259	96									
PB166779BL	PB166779BL	1304	95									
PB166779BS	PB166779BS	1308	91									
Q1383-02	OU4-PCS-TC-1	1315	94									
Q1383-04	OU4-PCS-TC-1	1318	94									
Q1383-06	OU4-PCS-TC-1	1321	96									
Q1383-08	OU4-PCS-TC-1	1324	94									
Q1383-10	OU4-PCS-TC-1	1328	94									
Q1383-12	OU4-PCS-TC-1	1331	94									
CCV02	CCV02	1334	84									
CCB02	CCB02	1341	97									
Q1383-16	OU4-PCS-TC-1	1355	93									
Q1383-18	OU4-PCS-TC-1	1358	96									
Q1383-20	OU4-PCS-TC-2	1401	94									
Q1383-14	OU4-PCS-TC-1	1405	93									
Q1383-22	OU4-CF-15-02	1408	92									
Q1383-22DUP	OU4-CF-15-02	1411	92									
Q1383-22L	OU4-CF-15-02	1414	93									
Q1383-22MS	OU4-CF-15-02	1418	92									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q1383

Sas No.: Q1383 SDG No.: Q1383

Instrument ID: P8

Start Date : 02/24/2025

Run Number: LB134802

End Date : 02/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1421	83									
CCB03	CCB03	1426	92									
Q1383-22MSD	OU4-CF-15-02	1430	91									
Q1383-22A	OU4-CF-15-02	1433	91									
PB166779TB	PB166779TB	1443	92									
CCV04	CCV04	1451	78									
CCB04	CCB04	1457	90									